

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082020\  
 Data File : VN063055.D  
 Acq On : 21 Aug 2020 00:03  
 Operator : JC/MD  
 Sample : L3698-06  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 35 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 S39-TB-2020-2

Quant Time: Aug 21 07:07:18 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N081820W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Aug 19 06:09:27 2020  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	144031	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	239648	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	245859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88529	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	96967	56.80	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	113.60%	
35) Dibromofluoromethane	7.55	113	83238	53.31	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	106.62%	
50) Toluene-d8	10.06	98	309150	51.54	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.08%	
62) 4-Bromofluorobenzene	12.38	95	102625	48.49	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.98%	

Target Compounds Qvalue

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082020\  
 Data File : VN063055.D  
 Acq On : 21 Aug 2020 00:03  
 Operator : JC/MD  
 Sample : L3698-06  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 35 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 S39-TB-2020-2

Quant Time: Aug 21 07:07:18 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N081820W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Aug 19 06:09:27 2020  
 Response via : Initial Calibration

