

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
 Data File : VN068240.D
 Acq On : 20 Aug 2021 13:08
 Operator : JC/MD
 Sample : M3428-07
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-03-24.5-25.0

Quant Time: Aug 20 15:17:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	414191	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.965	114	678301	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	659384	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	305980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	231582	52.365	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.740%	
35) Dibromofluoromethane	8.016	113	211764	49.727	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.460%	
50) Toluene-d8	10.443	98	816340	50.309	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.620%	
62) 4-Bromofluorobenzene	12.733	95	276258	51.142	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.280%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
Data File : VN068240.D
Acq On : 20 Aug 2021 13:08
Operator : JC/MD
Sample : M3428-07
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-03-24.5-25.0

Quant Time: Aug 20 15:17:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

