

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
 Data File : VN068248.D
 Acq On : 20 Aug 2021 16:27
 Operator : JC/MD
 Sample : PB138586ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138586ZHE#01

Manual Integrations
 APPROVED

MMDadoda
 8/24/2021 6:17:25 PM

Quant Time: Aug 20 23:59:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	420908	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	723025	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	768959	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	287034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	243152	54.104	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.200%			
35) Dibromofluoromethane	8.024	113	217453	47.904	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 95.800%			
50) Toluene-d8	10.443	98	893817	51.676	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.360%			
62) 4-Bromofluorobenzene	12.736	95	290069	50.377	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.760%			
Target Compounds					Qvalue	
16) Acetone	4.299	43	25503	19.436	ug/l #	88
20) Methylene Chloride	5.090	84	22422m	4.230	ug/l	
43) Isopropyl Acetate	8.566	43	37321	4.627	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
Data File : VN068248.D
Acq On : 20 Aug 2021 16:27
Operator : JC/MD
Sample : PB138586ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138586ZHE#01

Manual Integrations
APPROVED

MMDadoda
8/24/2021 6:17:25 PM

Quant Time: Aug 20 23:59:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

