

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
 Data File : VN068249.D
 Acq On : 20 Aug 2021 16:51
 Operator : JC/MD
 Sample : PB138586ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138586ZHE#02

Quant Time: Aug 20 23:59:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	406185	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	698734	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	748107	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	280220	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	230807	53.219	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	106.440%	
35) Dibromofluoromethane	8.024	113	211931	48.311	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	96.620%	
50) Toluene-d8	10.446	98	865520	51.780	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	103.560%	
62) 4-Bromofluorobenzene	12.733	95	282731	50.810	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	101.620%	
Target Compounds						
						Qvalue
16) Acetone	4.296	43	26204	20.694	ug/l	# 79
20) Methylene Chloride	5.095	84	21043	4.052	ug/l	# 83
43) Isopropyl Acetate	8.566	43	37861	4.857	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
Data File : VN068249.D
Acq On : 20 Aug 2021 16:51
Operator : JC/MD
Sample : PB138586ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138586ZHE#02

Quant Time: Aug 20 23:59:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

