

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
 Data File : VN068250.D
 Acq On : 20 Aug 2021 17:16
 Operator : JC/MD
 Sample : M3468-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-11-(647.00)

Quant Time: Aug 20 23:59:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	430905	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	755514	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	788211	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	293845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	256604	55.773	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.540%
35) Dibromofluoromethane	8.024	113	233713	49.272	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.540%
50) Toluene-d8	10.446	98	931950	51.564	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.120%
62) 4-Bromofluorobenzene	12.736	95	303724	50.481	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.960%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	30834	22.954	ug/l	# 90
20) Methylene Chloride	5.098	84	28197	5.703	ug/l	# 79
43) Isopropyl Acetate	8.566	43	31174	3.699	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082021\
Data File : VN068250.D
Acq On : 20 Aug 2021 17:16
Operator : JC/MD
Sample : M3468-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-11-(647.00)

Quant Time: Aug 20 23:59:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

