

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082024\
 Data File : VN083392.D
 Acq On : 20 Aug 2024 11:24
 Operator : JC\MD
 Sample : VN0820MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0820MBL01

Quant Time: Aug 21 02:22:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	112452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	226743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	113267	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	101887	63.654	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	127.300%#
35) Dibromofluoromethane	8.165	113	83284	58.846	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	117.700%
50) Toluene-d8	10.565	98	305648	57.895	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	115.800%#
62) 4-Bromofluorobenzene	12.847	95	130275	63.296	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	126.600%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082024\
Data File : VN083392.D
Acq On : 20 Aug 2024 11:24
Operator : JC\MD
Sample : VN0820MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0820MBL01

Quant Time: Aug 21 02:22:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

