

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082024\
 Data File : VN083393.D
 Acq On : 20 Aug 2024 11:46
 Operator : JC\MD
 Sample : VN0820WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0820WBL01

Quant Time: Aug 21 02:23:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	260624	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107256	58.732	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.460%	
35) Dibromofluoromethane	8.165	113	86658	53.270	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.540%	
50) Toluene-d8	10.565	98	307110	50.610	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.220%	
62) 4-Bromofluorobenzene	12.847	95	141728	59.909	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 119.820%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082024\
Data File : VN083393.D
Acq On : 20 Aug 2024 11:46
Operator : JC\MD
Sample : VN0820WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0820WBL01

Quant Time: Aug 21 02:23:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

