

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082024\
 Data File : VN083406.D
 Acq On : 20 Aug 2024 17:40
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Aug 21 02:26:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257996	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	280331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	139092	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109754	62.124	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	124.240%
35) Dibromofluoromethane	8.165	113	85926	53.358	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.720%
50) Toluene-d8	10.565	98	312550	52.031	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.060%
62) 4-Bromofluorobenzene	12.847	95	146209	62.432	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	124.860%#


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082024\
Data File : VN083406.D
Acq On : 20 Aug 2024 17:40
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Aug 21 02:26:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083406.D
Acq: 20 Aug 2024 17:40




Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Tgt Ion:168 Resp: 124118
Ion Ratio Lower Upper
168 100
99 70.5 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 62.124 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.005 min
Lab File: VN083406.D
Acq: 20 Aug 2024 17:40

Tgt Ion: 65 Resp: 109754
Ion Ratio Lower Upper
65 100
67 50.5 0.0 104.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN083406.D

Acq: 20 Aug 2024 17:40

Instrument :

MSVOA_N

ClientSampleId :

VIBLK





Tgt Ion:114 Resp: 257996

Ion Ratio Lower Upper

114 100

63 23.0 0.0 44.6

88 16.7 0.0 31.4



#35

Dibromofluoromethane

Concen: 53.358 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN083406.D

Acq: 20 Aug 2024 17:40




Tgt Ion:113 Resp: 85926


Ion Ratio Lower Upper

113 100

111 104.5 82.4 123.6

192 17.4 14.9 22.3









#50
Toluene-d8
Concen: 52.031 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN083406.D
Acq: 20 Aug 2024 17:40




Instrument :
MSVOA_N
ClientSampleId :
VIBLK


Tgt Ion: 98 Resp: 312550
Ion Ratio Lower Upper
98 100
100 65.1 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 62.432 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN083406.D
Acq: 20 Aug 2024 17:40

Tgt Ion: 95 Resp: 146209
Ion Ratio Lower Upper
95 100
174 69.2 0.0 159.2
176 66.9 0.0 147.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN083406.D
Acq: 20 Aug 2024 17:40

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Tgt Ion:117 Resp: 280331
Ion Ratio Lower Upper
117 100
82 61.7 47.5 71.3
119 32.2 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN083406.D
Acq: 20 Aug 2024 17:40

Tgt Ion:152 Resp: 139092
Ion Ratio Lower Upper
152 100
115 64.1 30.6 91.6
150 158.7 0.0 348.6

