

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082118\
 Data File : VN050800.D
 Acq On : 21 Aug 2018 23:25
 Operator : MD\SY
 Sample : PB112227ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112227ZHE#06

Quant Time: Aug 22 07:44:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	635636	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1037799	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	917136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	311466	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	418792	54.73	ug/l	0.00
Spiked Amount			Recovery	=	109.46%	
35) Dibromofluoromethane	7.59	113	394074	47.73	ug/l	0.00
Spiked Amount			Recovery	=	95.46%	
50) Toluene-d8	10.09	98	1438650	46.61	ug/l	0.00
Spiked Amount			Recovery	=	93.22%	
62) 4-Bromofluorobenzene	12.40	95	396087	38.61	ug/l	0.00
Spiked Amount			Recovery	=	77.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	19843	4.00	ug/l	99
18) Methyl Acetate	4.33	43	13388	1.85	ug/l	100
20) Methylene Chloride	4.55	84	34674	4.02	ug/l	94
43) Isopropyl Acetate	8.17	43	311041	25.72	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082118\
Data File : VN050800.D
Acq On : 21 Aug 2018 23:25
Operator : MD\SY
Sample : PB112227ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112227ZHE#06

Quant Time: Aug 22 07:44:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

