

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082118\
 Data File : VN050802.D
 Acq On : 22 Aug 2018 00:15
 Operator : MD\SY
 Sample : PB112227ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112227ZHE#08

Quant Time: Aug 22 07:56:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	597807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1007507	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	879678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	295760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	401497	55.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.58%	
35) Dibromofluoromethane	7.59	113	379414	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
50) Toluene-d8	10.09	98	1382422	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
62) 4-Bromofluorobenzene	12.40	95	380411	38.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.40%	
Target Compounds						
16) Acetone	3.82	43	20124	4.88	ug/l	99
18) Methyl Acetate	4.33	43	13522	1.99	ug/l	98
20) Methylene Chloride	4.56	84	31147	3.84	ug/l	96
43) Isopropyl Acetate	8.17	43	316215	26.93	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082118\
Data File : VN050802.D
Acq On : 22 Aug 2018 00:15
Operator : MD\SY
Sample : PB112227ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112227ZHE#08

Quant Time: Aug 22 07:56:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

