

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082118\
 Data File : VN050803.D
 Acq On : 22 Aug 2018 00:39
 Operator : MD\SY
 Sample : PB112227ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112227ZHE#09

Quant Time: Aug 22 08:01:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	623699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1046901	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	886027	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	274310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	422201	56.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.46%	
35) Dibromofluoromethane	7.59	113	393738	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
50) Toluene-d8	10.09	98	1431910	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
62) 4-Bromofluorobenzene	12.40	95	367053	35.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.94%	
Target Compounds						
16) Acetone	3.82	43	21961	5.44	ug/l	96
18) Methyl Acetate	4.34	43	14668	2.07	ug/l	96
20) Methylene Chloride	4.56	84	32242	3.81	ug/l	98
43) Isopropyl Acetate	8.17	43	333984	27.37	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082118\
Data File : VN050803.D
Acq On : 22 Aug 2018 00:39
Operator : MD\SY
Sample : PB112227ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112227ZHE#09

Quant Time: Aug 22 08:01:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

