

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082118\
 Data File : VN050804.D
 Acq On : 22 Aug 2018 1:04
 Operator : MD\SY
 Sample : PB112227ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112227ZHE#10

Quant Time: Aug 22 08:04:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	612409	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1021539	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	854309	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	253740	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	408058	55.35	ug/l	0.00
Spiked Amount			Recovery	=	110.70%	
35) Dibromofluoromethane	7.59	113	387148	47.64	ug/l	0.00
Spiked Amount			Recovery	=	95.28%	
50) Toluene-d8	10.09	98	1388741	45.71	ug/l	0.00
Spiked Amount			Recovery	=	91.42%	
62) 4-Bromofluorobenzene	12.40	95	350007	34.67	ug/l	0.00
Spiked Amount			Recovery	=	69.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	20988	5.10	ug/l	96
18) Methyl Acetate	4.34	43	15836	2.27	ug/l #	86
20) Methylene Chloride	4.56	84	32372	3.90	ug/l	97
43) Isopropyl Acetate	8.17	43	326882	27.46	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082118\
Data File : VN050804.D
Acq On : 22 Aug 2018 1:04
Operator : MD\SY
Sample : PB112227ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112227ZHE#10

Quant Time: Aug 22 08:04:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

