

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082118\
 Data File : VN050807.D
 Acq On : 22 Aug 2018 2:17
 Operator : MD\SY
 Sample : PB112227ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112227ZHE#13

Quant Time: Aug 22 08:16:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	610689	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1028929	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	868737	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	303452	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	412666	56.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.26%	
35) Dibromofluoromethane	7.59	113	393510	48.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.14%	
50) Toluene-d8	10.09	98	1378225	45.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.06%	
62) 4-Bromofluorobenzene	12.40	95	377057	37.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	74.16%	

Target Compounds						Qvalue
16) Acetone	3.83	43	19223	4.09	ug/l	98
18) Methyl Acetate	4.34	43	16715	2.41	ug/l #	77
20) Methylene Chloride	4.55	84	30504	3.68	ug/l	98
43) Isopropyl Acetate	8.17	43	324521	27.06	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082118\
Data File : VN050807.D
Acq On : 22 Aug 2018 2:17
Operator : MD\SY
Sample : PB112227ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112227ZHE#13

Quant Time: Aug 22 08:16:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

