

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082218\
 Data File : VN050818.D
 Acq On : 22 Aug 2018 14:38
 Operator : MD\SY
 Sample : PB112269TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112269TB

Quant Time: Aug 23 01:48:45 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	709294	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1130736	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	977595	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	393196	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	450465	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
35) Dibromofluoromethane	7.59	113	425518	47.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.60%	
50) Toluene-d8	10.09	98	1561338	46.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.84%	
62) 4-Bromofluorobenzene	12.40	95	481910	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
Target Compounds						
16) Acetone	3.82	43	21824	3.84	ug/l	96
18) Methyl Acetate	4.34	43	10458	1.30	ug/l	96
20) Methylene Chloride	4.55	84	55081	5.72	ug/l	97
43) Isopropyl Acetate	8.17	43	291371	22.11	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	37320	1.41	ug/l	98
95) Naphthalene	15.14	128	11661	6.06	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082218\
Data File : VN050818.D
Acq On : 22 Aug 2018 14:38
Operator : MD\SY
Sample : PB112269TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112269TB

Quant Time: Aug 23 01:48:45 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

