

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082218\
 Data File : VN050830.D
 Acq On : 22 Aug 2018 19:34
 Operator : MD\SY
 Sample : J4283-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-082118

Manual Integrations
 APPROVED

MMDadoda
 8/23/2018 12:27:21 PM

Quant Time: Aug 23 03:28:57 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	701962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1145686	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	958515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	356606	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	454190	53.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.50%	
35) Dibromofluoromethane	7.59	113	428791	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
50) Toluene-d8	10.09	98	1545672	45.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.72%	
62) 4-Bromofluorobenzene	12.40	95	430163	37.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.98%	
Target Compounds						
16) Acetone	3.83	43	25472	5.83	ug/l	98
18) Methyl Acetate	4.33	43	19183	2.40	ug/l #	71
20) Methylene Chloride	4.55	84	4793002	503.23	ug/l	98
29) Tetrahydrofuran	7.21	42	2544m	1.30	ug/l	
43) Isopropyl Acetate	8.17	43	354713	26.56	ug/l #	91
95) Naphthalene	15.14	128	43529	8.13	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082218\
Data File : VN050830.D
Acq On : 22 Aug 2018 19:34
Operator : MD\SY
Sample : J4283-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-082118

Manual Integrations
APPROVED

MMDadoda
8/23/2018 12:27:21 PM

Quant Time: Aug 23 03:28:57 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

