

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082218\
 Data File : VN050832.D
 Acq On : 22 Aug 2018 20:23
 Operator : MD\SY
 Sample : J4278-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-01-082118-B

Quant Time: Aug 23 03:33:20 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	683465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1127126	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	951094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	370230	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	438934	53.35	ug/l	0.00
Spiked Amount			Recovery	=	106.70%	
35) Dibromofluoromethane	7.59	113	413519	46.12	ug/l	0.00
Spiked Amount			Recovery	=	92.24%	
50) Toluene-d8	10.09	98	1522262	45.41	ug/l	0.00
Spiked Amount			Recovery	=	90.82%	
62) 4-Bromofluorobenzene	12.40	95	434507	39.00	ug/l	0.00
Spiked Amount			Recovery	=	78.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	21632	4.16	ug/l	99
18) Methyl Acetate	4.34	43	17602	2.27	ug/l	98
20) Methylene Chloride	4.55	84	49826	5.37	ug/l	97
29) Tetrahydrofuran	7.22	42	2241	1.18	ug/l #	75
30) Chloroform	7.38	83	23098	1.52	ug/l	97
43) Isopropyl Acetate	8.17	43	366778	27.92	ug/l #	90
95) Naphthalene	15.14	128	684880	46.70	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082218\
Data File : VN050832.D
Acq On : 22 Aug 2018 20:23
Operator : MD\SY
Sample : J4278-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-01-082118-B

Quant Time: Aug 23 03:33:20 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

