

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082219\
 Data File : VN057447.D
 Acq On : 22 Aug 2019 21:41
 Operator : JC/SP
 Sample : K4403-28
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-H4-(2.5)

Quant Time: Aug 23 05:32:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	454575	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	795497	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	657866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	166402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	375814	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
35) Dibromofluoromethane	7.57	113	256662	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
50) Toluene-d8	10.08	98	1000536	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
62) 4-Bromofluorobenzene	12.40	95	277403	40.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.58%	
Target Compounds						
16) Acetone	3.79	43	23968	6.527	ug/l	90
20) Methylene Chloride	4.53	84	12925	2.302	ug/l	89
43) Isopropyl Acetate	8.16	43	99184	6.735	ug/l	100
95) Naphthalene	15.13	128	7786	5.413	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082219\
Data File : VN057447.D
Acq On : 22 Aug 2019 21:41
Operator : JC/SP
Sample : K4403-28
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-H4-(2.5)

Quant Time: Aug 23 05:32:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

