

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082219\
 Data File : VN057451.D
 Acq On : 22 Aug 2019 23:23
 Operator : JC/SP
 Sample : K4403-34
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-I-(2.5)

Quant Time: Aug 23 05:40:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	447702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	782931	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	625541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	145861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	368054	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
35) Dibromofluoromethane	7.57	113	256603	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
50) Toluene-d8	10.08	98	980762	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
62) 4-Bromofluorobenzene	12.40	95	256435	37.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.68%	
Target Compounds						
20) Methylene Chloride	4.52	84	10475	1.881	ug/l	89
43) Isopropyl Acetate	8.16	43	119672	8.256	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082219\
Data File : VN057451.D
Acq On : 22 Aug 2019 23:23
Operator : JC/SP
Sample : K4403-34
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-I-(2.5)

Quant Time: Aug 23 05:40:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

