

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082219\
 Data File : VN057452.D
 Acq On : 22 Aug 2019 23:49
 Operator : JC/SP
 Sample : K4403-36
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-J(0)

Quant Time: Aug 23 05:42:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	436968	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	768642	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	637314	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	160191	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	365301	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
35) Dibromofluoromethane	7.57	113	244429	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
50) Toluene-d8	10.09	98	977128	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
62) 4-Bromofluorobenzene	12.40	95	270754	40.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.40%	
Target Compounds						
20) Methylene Chloride	4.53	84	11267	2.081	ug/l	96
43) Isopropyl Acetate	8.16	43	116938	8.217	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082219\
Data File : VN057452.D
Acq On : 22 Aug 2019 23:49
Operator : JC/SP
Sample : K4403-36
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-J-(0)

Quant Time: Aug 23 05:42:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

