


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN082219\
Data File : VN057436.D
Acq On : 22 Aug 2019 16:05
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Title : SW846 8260
Last Update : Fri Aug 23 05:07:29 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3292

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.7	12981	PASS
75	95	30	60	57.0	32576	PASS
95	95	100	100	100.0	57157	PASS
96	95	5	9	6.6	3755	PASS
173	174	0.00	2	0.5	194	PASS
174	95	50	100	66.0	37712	PASS
175	174	5	9	8.6	3258	PASS
176	174	95	101	96.4	36344	PASS
177	176	5	9	6.2	2266	PASS