

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
 Data File : VN078982.D
 Acq On : 22 Aug 2023 10:22
 Operator : JC\MD
 Sample : VN0822WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0822WBL01

Quant Time: Aug 23 02:57:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	178718	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	333633	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	300780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89415	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	145083	52.169	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.340%
35) Dibromofluoromethane	8.171	113	120261	53.548	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.100%
50) Toluene-d8	10.571	98	405328	48.606	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.220%
62) 4-Bromofluorobenzene	12.853	95	121491	44.573	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
Data File : VN078982.D
Acq On : 22 Aug 2023 10:22
Operator : JC\MD
Sample : VN0822WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0822WBL01

Quant Time: Aug 23 02:57:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

