

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
 Data File : VN078992.D
 Acq On : 22 Aug 2023 14:23
 Operator : JC\MD
 Sample : PB154968ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154968ZHE#07

Quant Time: Aug 23 03:02:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	174989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	353997	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	342312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	105827	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	143400	52.663	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.320%	
35) Dibromofluoromethane	8.171	113	118911	49.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.800%	
50) Toluene-d8	10.571	98	444445	50.231	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.460%	
62) 4-Bromofluorobenzene	12.853	95	137479	47.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.080%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	28726	25.567	ug/l	98
20) Methylene Chloride	5.289	84	7913	3.305	ug/l	91
25) 2-Butanone	7.500	43	17333	9.622	ug/l #	67
43) Isopropyl Acetate	8.694	43	200106	28.229	ug/l #	72
51) 4-Methyl-2-Pentanone	10.453	43	64915	15.440	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
Data File : VN078992.D
Acq On : 22 Aug 2023 14:23
Operator : JC\MD
Sample : PB154968ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154968ZHE#07

Quant Time: Aug 23 03:02:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

