

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
 Data File : VN078995.D
 Acq On : 22 Aug 2023 15:35
 Operator : JC\MD
 Sample : PB154968ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154968ZHE#10

Quant Time: Aug 23 03:03:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	173957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	344195	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	334651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	145312	53.682	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.360%
35) Dibromofluoromethane	8.171	113	118783	51.267	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.540%
50) Toluene-d8	10.571	98	435074	50.572	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.140%
62) 4-Bromofluorobenzene	12.853	95	135225	48.089	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.180%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	25613	22.932	ug/l #	76
20) Methylene Chloride	5.277	84	11284	4.741	ug/l #	76
25) 2-Butanone	7.506	43	10703	5.977	ug/l	94
43) Isopropyl Acetate	8.694	43	229763	33.336	ug/l #	71
51) 4-Methyl-2-Pentanone	10.453	43	42912	10.497	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
Data File : VN078995.D
Acq On : 22 Aug 2023 15:35
Operator : JC\MD
Sample : PB154968ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154968ZHE#10

Quant Time: Aug 23 03:03:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

