

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
 Data File : VN078999.D
 Acq On : 22 Aug 2023 17:12
 Operator : JC\MD
 Sample : 04060-01DL 5X
 Misc : 5.07g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1002DL

Quant Time: Aug 22 17:46:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	187144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	349125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	325552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	144550	49.637	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.280%
35) Dibromofluoromethane	8.171	113	117330	49.925	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.840%
50) Toluene-d8	10.571	98	409804	46.962	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.920%
62) 4-Bromofluorobenzene	12.853	95	143666	50.369	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.740%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	111063	92.430	ug/l #	86
52) Toluene	10.635	92	1518873	236.753	ug/l	98
67) Ethyl Benzene	11.970	91	60639	4.812	ug/l	99
68) m/p-Xylenes	12.070	106	59739	12.679	ug/l	95
69) o-Xylene	12.400	106	9845	2.148	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082223\
Data File : VN078999.D
Acq On : 22 Aug 2023 17:12
Operator : JC\MD
Sample : 04060-01DL 5X
Misc : 5.07g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1002DL

Quant Time: Aug 22 17:46:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

