

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082225\
 Data File : VN087630.D
 Acq On : 22 Aug 2025 10:47
 Operator : JC\MD
 Sample : VN0822WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0822WBL01

Quant Time: Aug 25 01:16:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	212794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	479771	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	448296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	206141	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	234010	53.673	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.340%
35) Dibromofluoromethane	8.147	113	171259	49.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	10.547	98	624323	48.155	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.300%
62) 4-Bromofluorobenzene	12.823	95	219804	47.672	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082225\
Data File : VN087630.D
Acq On : 22 Aug 2025 10:47
Operator : JC\MD
Sample : VN0822WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0822WBL01

Quant Time: Aug 25 01:16:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

