

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082225\
 Data File : VN087642.D
 Acq On : 22 Aug 2025 15:26
 Operator : JC\MD
 Sample : Q2936-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-19C-082125

Quant Time: Aug 25 01:21:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	178938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	421488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	403205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	189324	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	210412	57.392	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.780%
35) Dibromofluoromethane	8.147	113	151482	49.778	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.560%
50) Toluene-d8	10.547	98	545197	47.867	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.740%
62) 4-Bromofluorobenzene	12.829	95	194723	48.073	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.140%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	9885	4.953	ug/l	95
27) cis-1,2-Dichloroethene	7.471	96	58449	17.486	ug/l	94
40) Benzene	8.588	78	6354	0.444	ug/l	95
44) Trichloroethene	9.335	130	85049	26.015	ug/l	100
64) Tetrachloroethene	11.082	164	61116	21.847	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082225\
Data File : VN087642.D
Acq On : 22 Aug 2025 15:26
Operator : JC\MD
Sample : Q2936-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-19C-082125

Quant Time: Aug 25 01:21:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

