

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082225\
 Data File : VN087638.D
 Acq On : 22 Aug 2025 14:01
 Operator : JC\MD
 Sample : Q2936-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Aug 25 01:20:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	166980	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	383570	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	362013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	167155	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	182148	53.241	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.480%
35) Dibromofluoromethane	8.153	113	133050	48.043	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.080%
50) Toluene-d8	10.547	98	493240	47.586	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.180%
62) 4-Bromofluorobenzene	12.829	95	170572	46.273	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.540%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082225\
Data File : VN087638.D
Acq On : 22 Aug 2025 14:01
Operator : JC\MD
Sample : Q2936-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Aug 25 01:20:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.212 min Scan# 1064
Delta R.T. 0.006 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
TB


Tgt Ion:168 Resp: 166980
Ion Ratio Lower Upper
168 100
99 64.8 57.2 85.8



#33
1,2-Dichloroethane-d4
Concen: 53.241 ug/l
RT: 8.565 min Scan# 1124
Delta R.T. 0.006 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01

Tgt Ion: 65 Resp: 182148
Ion Ratio Lower Upper
65 100
67 49.4 0.0 100.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.083 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
TB


Tgt Ion:114 Resp: 383570
Ion Ratio Lower Upper
114 100
63 23.0 0.0 45.2
88 16.7 0.0 33.4



#35
Dibromofluoromethane
Concen: 48.043 ug/l
RT: 8.153 min Scan# 1054
Delta R.T. -0.000 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01

Tgt Ion:113 Resp: 133050
Ion Ratio Lower Upper
113 100
111 105.8 82.8 124.2
192 16.4 12.8 19.2









#50
Toluene-d8
Concen: 47.586 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01




Instrument :
MSVOA_N
ClientSampleId :
TB


Tgt Ion: 98 Resp: 493240
Ion Ratio Lower Upper
98 100
100 63.8 52.8 79.2



#62
4-Bromofluorobenzene
Concen: 46.273 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.006 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01

Tgt Ion: 95 Resp: 170572
Ion Ratio Lower Upper
95 100
174 70.8 0.0 138.4
176 66.9 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.841 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01

Instrument :
MSVOA_N
ClientSampleId :
TB

Tgt Ion:117 Resp: 362013
Ion Ratio Lower Upper
117 100
82 58.7 47.4 71.2
119 31.4 24.3 36.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.006 min
Lab File: VN087638.D
Acq: 22 Aug 2025 14:01

Tgt Ion:152 Resp: 167155
Ion Ratio Lower Upper
152 100
115 65.7 32.3 96.8
150 158.7 0.0 351.0

