

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082318\
 Data File : VN050847.D
 Acq On : 23 Aug 2018 12:43
 Operator : MD\SY
 Sample : PB112301ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Aug 24 01:09:44 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	671700	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1111160	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	926878	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	314246	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	440129	54.43	ug/l	0.00
Spiked Amount			Recovery	=	108.86%	
35) Dibromofluoromethane	7.59	113	421949	47.73	ug/l	0.00
Spiked Amount			Recovery	=	95.46%	
50) Toluene-d8	10.09	98	1498021	45.33	ug/l	0.00
Spiked Amount			Recovery	=	90.66%	
62) 4-Bromofluorobenzene	12.40	95	398077	36.25	ug/l	0.00
Spiked Amount			Recovery	=	72.50%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.06	50	2750	0.25	ug/l	# 79
5) Bromomethane	2.58	94	2848	Below	Cal	# 99
10) Methyl Iodide	3.95	142	529	6.40	ug/l	# 69
16) Acetone	3.82	43	16754	1.73	ug/l	# 97
18) Methyl Acetate	4.34	43	11876	1.56	ug/l	# 69
20) Methylene Chloride	4.55	84	22201	2.44	ug/l	# 92
25) 2-Butanone	6.85	43	3930	1.21	ug/l	# 92
29) Tetrahydrofuran	7.22	42	2370	1.27	ug/l	# 94
31) Cyclohexane	7.67	56	11297	1.28	ug/l	# 1
36) 1,1-Dichloropropene	7.67	75	45625	3.48	ug/l	# 49
37) Ethyl Acetate	6.85	43	3930	0.52	ug/l	# 67
38) Carbon Tetrachloride	7.67	117	59665	4.50	ug/l	# 17
41) Methacrylonitrile	7.22	41	1249	0.31	ug/l	# 36
43) Isopropyl Acetate	8.17	43	298199	23.03	ug/l	# 95
48) Methyl methacrylate	9.18	41	7796	1.18	ug/l	# 19
51) 4-Methyl-2-Pentanone	9.99	43	1670	0.22	ug/l	# 95
56) Ethyl methacrylate	10.22	69	2160	2.53	ug/l	# 64
59) 2-Hexanone	10.77	43	16719	3.25	ug/l	# 59
70) Styrene	11.96	104	161	1.28	ug/l	# 26
71) Bromoform	12.41	173	1605	0.26	ug/l	# 1
75) 1,1,2,2-Tetrachloroethane	12.40	83	173	Below	Cal	# 1
76) 1,2,3-Trichloropropane	12.40	75	189650	34.46	ug/l	# 100
81) trans-1,4-Dichloro-2-buten	12.40	75	189650	116.48	ug/l	# 15
93) 1,2,4-Trichlorobenzene	14.92	180	62	4.09	ug/l	# 12
95) Naphthalene	15.14	128	393	5.43	ug/l	# 69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082318\
Data File : VN050847.D
Acq On : 23 Aug 2018 12:43
Operator : MD\SY
Sample : PB112301ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Aug 24 01:09:44 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.00 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43


Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:168 Resp: 671700

Ion Ratio Lower Upper

168 100

99 48.8 39.6 59.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67



300000

200000

100000

0

Time-->



#3

Chloromethane

Concen: 0.25 ug/l

RT: 2.06 min Scan# 145

Delta R.T. -0.00 min

Lab File: VN050847.D


Acq: 23 Aug 2018 12:43

Tgt Ion: 50 Resp: 2750

Ion Ratio Lower Upper

50 100

52 21.3 26.3 39.5#



Abundance Ion 49.90 (49.60 to 50.60): VN

Ion 51.90 (51.60 to 52.60): VN

2500

2000


1500


1000

500

0

Time-->





#5

Bromomethane

Concen: Below Cal

RT: 2.58 min Scan# 307

Delta R.T. 0.02 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43

Instrument :

MSVOA_N


ClientSampled :

Tot Ion: 94 Resp: 2848




Ion Ratio Lower Upper

94 100

96 93.9 76.2 114.2



Abundance Ion 93.90 (93.60 to 94.60): VN050847.D (-209) (-)



#10

Methyl Iodide

Concen: 6.40 ug/l

RT: 3.95 min Scan# 734

Delta R.T. -0.00 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43


Tgt Ion: 142 Resp: 529

Ion Ratio Lower Upper



142 100


127 21.0 31.7 47.5#

141 0.0 11.4 17.0#



Abundance Ion 141.80 (141.50 to 142.50): VN050847.D (-642) (-)





#16

Acetone

Concen: 1.73 ug/l

RT: 3.82 min Scan# 695

Delta R.T. 0.01 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 43 Resp: 16754

Ion Ratio Lower Upper

43 100

58 32.4 27.2 40.8



#18

Methyl Acetate

Concen: 1.56 ug/l

RT: 4.34 min Scan# 855

Delta R.T. 0.01 min

Lab File: VN050847.D




Acq: 23 Aug 2018 12:43


Tgt Ion: 43 Resp: 11876

Ion Ratio Lower Upper

43 100

74 9.4 19.8 29.8#









#20
Methvlene Chloride
Concen: 2.44 ug/l
RT: 4.55 min Scan# 922
Delta R.T. 0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 84 Resp: 22201
Ion Ratio Lower Upper
84 100
49 103.4 93.2 139.8
51 32.8 29.0 43.6
86 60.7 49.8 74.8



#25
2-Butanone
Concen: 1.21 ug/l
RT: 6.85 min Scan# 1637
Delta R.T. 0.02 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion: 43 Resp: 3930
Ion Ratio Lower Upper
43 100
72 23.7 22.3 33.5





#29

Tetrahydrofuran

Concen: 1.27 ug/l

RT: 7.22 min Scan# 1750

Delta R.T. 0.00 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43

Instrument :
MSVOA_N
ClientSampled :





Tot Ion: 42 Resp: 2370

Ion Ratio Lower Upper

42 100

72 41.5 35.5 53.3

71 36.6 33.7 50.5



#31

Cyclohexane

Concen: 1.28 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.01 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43




Tgt Ion: 56 Resp: 11297


Ion Ratio Lower Upper

56 100

69 195.1 26.2 39.4#

84 149.0 68.3 102.5#





#33

1,2-Dichloroethane-d4

Concen: 54.43 ug/l

RT: 8.03 min Scan# 2003

Delta R.T. 0.00 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 65 Resp: 440129

Ion Ratio Lower Upper

65 100

67 54.3 0.0 110.4



Abundance Ion 64.90 (64.60 to 65.60): VN050782.D (-1983) (-)

Ion 66.90 (66.60 to 67.60): VN050782.D (-1983) (-)

8.03

200000

150000



100000

50000

0

7.90 8.00 8.10

Time-->



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.59 min Scan# 2177

Delta R.T. 0.00 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43


Tgt Ion:114 Resp: 1111160

Ion Ratio Lower Upper

114 100

63 19.6 0.0 38.6

88 15.6 0.0 30.6



Abundance Ion 113.90 (113.60 to 114.60): VN050782.D (-2158) (-)

Ion 63.00 (62.70 to 63.70): VN050782.D (-2158) (-)

Ion 87.90 (87.60 to 88.60): VN050782.D (-2158) (-)

8.59

600000


400000


200000

0

8.50 8.60 8.70 8.80

Time-->









#35
Dibromofluoromethane
Concen: 47.73 ug/l
RT: 7.59 min Scan# 1867
Delta R.T. -0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:113 Resp: 421949
Ion Ratio Lower Upper
113 100
111 101.7 82.6 123.8
192 22.8 18.0 27.0



#36
1,1-Dichloropropene
Concen: 3.48 ug/l
RT: 7.67 min Scan# 1890
Delta R.T. -0.13 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion: 75 Resp: 45625
Ion Ratio Lower Upper
75 100
110 9.2 18.1 54.4#
77 0.0 24.8 37.2#









#37
Ethyl Acetate
Concen: 0.52 ug/l
RT: 6.85 min Scan# 1637
Delta R.T. -0.08 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 43 Resp: 3930
Ion Ratio Lower Upper
43 100
61 0.0 12.0 18.0#
70 0.0 8.6 13.0#



#38
Carbon Tetrachloride
Concen: 4.50 ug/l
RT: 7.67 min Scan# 1890
Delta R.T. -0.11 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion: 117 Resp: 59665
Ion Ratio Lower Upper
117 100
119 6.6 76.3 114.5#
121 0.0 24.6 36.8#









#41
Methacrylonitrile
Concen: 0.31 ug/l
RT: 7.22 min Scan# 1752
Delta R.T. 0.05 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 41 Resp: 1249
Ion Ratio Lower Upper
41 100
39 70.5 45.4 68.0#
67 0.0 70.6 106.0#
52 0.0 27.1 40.7#



#43
Isopropyl Acetate
Concen: 23.03 ug/l
RT: 8.17 min Scan# 2046
Delta R.T. 0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion: 43 Resp: 298199
Ion Ratio Lower Upper
43 100
61 19.0 17.5 26.3
87 12.3 11.3 16.9









#48
Methyl methacrylate
Concen: 1.18 ug/l
RT: 9.18 min Scan# 2361
Delta R.T. -0.02 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion: 41 Resp: 7796
Ion Ratio Lower Upper
41 100
69 0.0 74.7 112.1#
39 16.2 42.2 63.2#




Abundance Ion 41.00 (40.70 to 41.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 39.00 (38.70 to 39.70): VN0






#50
Toluene-d8
Concen: 45.33 ug/l
RT: 10.09 min Scan# 2645
Delta R.T. 0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion: 98 Resp: 1498021
Ion Ratio Lower Upper
98 100
100 64.1 51.5 77.3



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0









#51
4-Methyl-2-Pentanone
Concen: 0.22 ug/l
RT: 9.99 min Scan# 2612
Delta R.T. 0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 1670
Ion Ratio Lower Upper
43 100
58 43.4 32.4 48.6



#56
Ethyl methacrylate
Concen: 2.53 ug/l
RT: 10.22 min Scan# 2685
Delta R.T. -0.20 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion: 69 Resp: 2160
Ion Ratio Lower Upper
69 100
41 33.1 48.3 72.5#
39 49.5 24.0 36.0#









#59
2-Hexanone
Concen: 3.25 ug/l
RT: 10.77 min Scan# 2854
Delta R.T. 0.01 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 43 Resp: 16719
Ion Ratio Lower Upper
43 100
58 26.5 28.3 85.0#



#62
4-Bromofluorobenzene
Concen: 36.25 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. 0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion: 95 Resp: 398077
Ion Ratio Lower Upper
95 100
174 92.2 0.0 183.0
176 88.8 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Instrument :
MSVOA_N
ClientSampleId :

Tot Ion:117 Resp: 926878
Ion Ratio Lower Upper
117 100
82 53.3 42.6 64.0
119 31.9 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VN0
Ion 118.90 (118.60 to 119.60): V






#70
Styrene
Concen: 1.28 ug/l
RT: 11.96 min Scan# 3226
Delta R.T. 0.00 min
Lab File: VN050847.D
Acq: 23 Aug 2018 12:43

Tgt Ion:104 Resp: 161
Ion Ratio Lower Upper
104 100
78 0.0 38.2 57.4#
103 0.0 44.4 66.6#



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VN0
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 0.26 ug/l

RT: 12.41 min Scan# 3364

Delta R.T. 0.28 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:173 Resp: 1605

Ion Ratio Lower Upper

173 100

175 1349.6 24.4 73.4#




254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43


Tgt Ion:152 Resp: 314246

Ion Ratio Lower Upper

152 100

115 55.9 27.8 83.4



150 157.4 0.0 351.4




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.40 min Scan# 3362

Delta R.T. -0.10 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 83 Resp: 173

Ion Ratio Lower Upper

83 100

131 217.9 5.3 15.9#




85 0.0 31.9 95.7#



Abundance Ion 82.90 (82.60 to 83.60): VNO

Ion 130.80 (130.50 to 131.50): VNO

Ion 84.90 (84.60 to 85.60): VNO



#76

1,2,3-Trichloropropane

Concen: 34.46 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN050847.D


Acq: 23 Aug 2018 12:43

Tgt Ion: 75 Resp: 189650

Ion Ratio Lower Upper



75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 77.00 (76.70 to 77.70): VNO





#81

trans-1,4-Dichloro-2-butene

Concen: 116.48 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 75 Resp: 189650

Ion Ratio Lower Upper

75 100

53 0.1 68.9 103.3#




89 0.0 36.1 54.1#



Abundance Ion 74.90 (74.60 to 75.60): VN050847.D (-3363) (-)

Ion 53.00 (52.70 to 53.70): VN050847.D (-3363) (-)

Ion 88.90 (88.60 to 89.60): VN050847.D (-3363) (-)



#93

1,2,4-Trichlorobenzene

Concen: 4.09 ug/l

RT: 14.92 min Scan# 4145

Delta R.T. 0.01 min

Lab File: VN050847.D

Acq: 23 Aug 2018 12:43


Tgt Ion: 180 Resp: 62

Ion Ratio Lower Upper

180 100

182 0.0 48.2 144.6#



145 0.0 14.1 42.2#




Abundance Ion 179.80 (179.50 to 180.50): VN050847.D (-4145) (-)

Ion 181.80 (181.50 to 182.50): VN050847.D (-4145) (-)

Ion 144.80 (144.50 to 145.50): VN050847.D (-4145) (-)





#95
 Naphthalene
 Concen: 5.43 ug/l
 RT: 15.14 min Scan# 4214
 Delta R.T. 0.01 min
 Lab File: VN050847.D
 Acq: 23 Aug 2018 12:43

Instrument :
 MSVOA_N
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
128	100		
127	0.0	10.5	15.7#
129	0.0	8.7	13.1#

