

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082318\
 Data File : VN050839.D
 Acq On : 23 Aug 2018 9:26
 Operator : MD\SY
 Sample : VN0823MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0823MBL01

Quant Time: Aug 24 01:32:58 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	662351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1053534	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	913447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	313812	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	426774	53.52	ug/l	0.00
Spiked Amount						
			Recovery	=	107.04%	
35) Dibromofluoromethane	7.59	113	401784	47.94	ug/l	0.00
Spiked Amount						
			Recovery	=	95.88%	
50) Toluene-d8	10.09	98	1475867	47.10	ug/l	0.00
Spiked Amount						
			Recovery	=	94.20%	
62) 4-Bromofluorobenzene	12.40	95	404974	38.89	ug/l	0.00
Spiked Amount						
			Recovery	=	77.78%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082318\
Data File : VN050839.D
Acq On : 23 Aug 2018 9:26
Operator : MD\SY
Sample : VN0823MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0823MBL01

Quant Time: Aug 24 01:32:58 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.00 min

Lab File: VN050839.D

Acq: 23 Aug 2018 9:26

Instrument :

MSVOA_N

ClientSampleId :


VN0823MBL01

Tot Ion:168 Resp: 662351

Ion Ratio Lower Upper

168 100

99 50.0 39.6 59.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

250000

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 53.52 ug/l

RT: 8.03 min Scan# 2002

Delta R.T. 0.00 min

Lab File: VN050839.D


Acq: 23 Aug 2018 9:26

Tgt Ion: 65 Resp: 426774

Ion Ratio Lower Upper

65 100

67 54.3 0.0 110.4



Abundance Ion 64.90 (64.60 to 65.60): VN

Ion 66.90 (66.60 to 67.60): VN

8.03

200000


150000


100000

50000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2176
 Delta R.T. 0.00 min
 Lab File: VN050839.D
 Acq: 23 Aug 2018 9:26




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0823MBL01


Tot Ion:114 Resp: 1053534
 Ion Ratio Lower Upper
 114 100
 63 19.5 0.0 38.6
 88 14.8 0.0 30.6



#35
 Dibromofluoromethane
 Concen: 47.94 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN050839.D
 Acq: 23 Aug 2018 9:26

Tgt Ion:113 Resp: 401784
 Ion Ratio Lower Upper
 113 100
 111 104.6 82.6 123.8
 192 22.8 18.0 27.0





#50

Toluene-d8

Concen: 47.10 ug/l

RT: 10.09 min Scan# 2644

Delta R.T. 0.00 min

Lab File: VN050839.D

Acq: 23 Aug 2018 9:26

Instrument :

MSVOA_N

ClientSampleId :





VN0823MBL01

Tot Ion: 98 Resp: 1475867

Ion Ratio Lower Upper

98 100

100 63.3 51.5 77.3



#62

4-Bromofluorobenzene

Concen: 38.89 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.00 min

Lab File: VN050839.D

Acq: 23 Aug 2018 9:26




Tgt Ion: 95 Resp: 404974


Ion Ratio Lower Upper

95 100

174 92.1 0.0 183.0

176 87.1 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3053
Delta R.T. 0.00 min
Lab File: VN050839.D
Acq: 23 Aug 2018 9:26

Instrument :
MSVOA_N
ClientSampleId :
VN0823MBL01

Tot Ion:117 Resp: 913447
Ion Ratio Lower Upper
117 100
82 53.9 42.6 64.0
119 32.5 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.35 min Scan# 3656
Delta R.T. 0.00 min
Lab File: VN050839.D
Acq: 23 Aug 2018 9:26

Tgt Ion:152 Resp: 313812
Ion Ratio Lower Upper
152 100
115 55.8 27.8 83.4
150 156.1 0.0 351.4

