

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082321\
 Data File : VN068260.D
 Acq On : 23 Aug 2021 10:54
 Operator : JC/MD
 Sample : VN0823MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0823MBL01

Quant Time: Aug 24 03:52:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	576030	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	943267	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	918041	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	322040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	272508	44.307	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.620%
35) Dibromofluoromethane	8.021	113	250313	42.268	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	84.540%
50) Toluene-d8	10.443	98	991145	43.924	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.840%
62) 4-Bromofluorobenzene	12.733	95	308758	41.103	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	82.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082321\
Data File : VN068260.D
Acq On : 23 Aug 2021 10:54
Operator : JC/MD
Sample : VN0823MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0823MBL01

Quant Time: Aug 24 03:52:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

