

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082321\
 Data File : VN068270.D
 Acq On : 23 Aug 2021 15:43
 Operator : JC/MD
 Sample : M3505-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRILL-CUTTING

Quant Time: Aug 24 03:54:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	490184	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	839004	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	894741	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	330526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	289116	55.240	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.480%
35) Dibromofluoromethane	8.024	113	257696	48.922	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.840%
50) Toluene-d8	10.443	98	1056653	52.646	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.300%
62) 4-Bromofluorobenzene	12.736	95	336438	50.353	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.700%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	15605	10.212	ug/l	91
20) Methylene Chloride	5.095	84	35746	6.610	ug/l #	83
43) Isopropyl Acetate	8.568	43	31346	3.349	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082321\
Data File : VN068270.D
Acq On : 23 Aug 2021 15:43
Operator : JC/MD
Sample : M3505-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRILL-CUTTING

Quant Time: Aug 24 03:54:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

