

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
 Data File : VN074065.D
 Acq On : 23 Aug 2022 13:58
 Operator : JC\MD
 Sample : N4279-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-050-IDWGW-20220819

Quant Time: Aug 24 01:20:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 27 18:07:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.586	128	66166	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.904	114	347042	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.686	117	296618	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.369	65	171078	29.904	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.667%
60) 4-Bromofluorobenzene	12.674	95	135231	25.748	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.833%
63) Toluene-d8	10.380	98	486083	29.103	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.000%
Target Compounds						
25) Chloroform	7.745	83	23953	2.616	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
Data File : VN074065.D
Acq On : 23 Aug 2022 13:58
Operator : JC\MD
Sample : N4279-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-050-IDWGW-20220819

Quant Time: Aug 24 01:20:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 27 18:07:26 2022
Response via : Initial Calibration

