

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
 Data File : VN074074.D
 Acq On : 23 Aug 2022 17:39
 Operator : JC\MD
 Sample : N4247-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-5

Quant Time: Aug 23 18:53:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	477464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	739764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	735116	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	360766	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	248142	53.358	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.720%	
35) Dibromofluoromethane	7.951	113	240610	54.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.000%	
50) Toluene-d8	10.381	98	815475	48.129	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.260%	
62) 4-Bromofluorobenzene	12.675	95	328617	58.439	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.880%	
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	73755	33.616	ug/l	98
20) Methylene Chloride	5.016	84	32809	7.087	ug/l	93
37) Ethyl Acetate	7.340	43	71386	9.697	ug/l	100
40) Benzene	8.392	78	49268	2.569	ug/l	98
43) Isopropyl Acetate	8.498	43	264900	25.259	ug/l	93
52) Toluene	10.445	92	43275	3.291	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
Data File : VN074074.D
Acq On : 23 Aug 2022 17:39
Operator : JC\MD
Sample : N4247-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-5

Quant Time: Aug 23 18:53:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

