

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
 Data File : VN074076.D
 Acq On : 23 Aug 2022 18:29
 Operator : JC\MD
 Sample : N4247-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-7

Quant Time: Aug 23 18:53:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	470056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	736141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	732223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	374981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	248065	54.182	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.360%
35) Dibromofluoromethane	7.951	113	237282	54.504	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.000%
50) Toluene-d8	10.381	98	806956	47.860	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.720%
62) 4-Bromofluorobenzene	12.675	95	332502	59.421	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.840%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	69646	32.244	ug/l	95
20) Methylene Chloride	5.010	84	34702	7.614	ug/l	97
40) Benzene	8.398	78	35681	1.870	ug/l	91
43) Isopropyl Acetate	8.498	43	284702	27.281	ug/l #	92
52) Toluene	10.445	92	30025	2.294	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
Data File : VN074076.D
Acq On : 23 Aug 2022 18:29
Operator : JC\MD
Sample : N4247-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-7

Quant Time: Aug 23 18:53:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

