

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
 Data File : VN074077.D
 Acq On : 23 Aug 2022 18:54
 Operator : JC\MD
 Sample : N4261-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C5

Quant Time: Aug 24 01:14:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	461457	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	728331	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	706033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	342784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	241398	53.709	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.420%	
35) Dibromofluoromethane	7.957	113	235518	54.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.360%	
50) Toluene-d8	10.380	98	790709	47.400	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.800%	
62) 4-Bromofluorobenzene	12.674	95	315075	56.910	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.820%	
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	54596	25.747	ug/l	93
20) Methylene Chloride	5.016	84	37473	8.375	ug/l	95
37) Ethyl Acetate	7.339	43	69486	9.587	ug/l	98
43) Isopropyl Acetate	8.498	43	278279	26.951	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082322\
Data File : VN074077.D
Acq On : 23 Aug 2022 18:54
Operator : JC\MD
Sample : N4261-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C5

Quant Time: Aug 24 01:14:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

