

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
 Data File : VN050862.D
 Acq On : 24 Aug 2018 8:41
 Operator : MD\SY
 Sample : VN0824WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824WBL01

Quant Time: Aug 24 14:12:41 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	106561	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	544624	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	464127	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	209249	28.89	ug/l	0.00
Spiked Amount 30.000	Range	50 - 169	Recovery	=	96.30%	
60) 4-Bromofluorobenzene	12.40	95	148591	20.84	ug/l	0.00
Spiked Amount 30.000	Range	56 - 143	Recovery	=	69.47%	
63) Toluene-d8	10.09	98	641670	28.94	ug/l	0.00
Spiked Amount 30.000	Range	66 - 137	Recovery	=	96.47%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
Data File : VN050862.D
Acq On : 24 Aug 2018 8:41
Operator : MD\SY
Sample : VN0824WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0824WBL01

Quant Time: Aug 24 14:12:41 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

