

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
 Data File : VN050864.D
 Acq On : 24 Aug 2018 9:37
 Operator : MD\SY
 Sample : J4635-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1

Quant Time: Aug 24 14:16:40 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	108931	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	543860	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	457063	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	214796	29.01	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.70%
60) 4-Bromofluorobenzene	12.40	95	148452	21.14	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	70.47%
63) Toluene-d8	10.09	98	640816	29.35	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.83%

Target Compounds					Qvalue
25) Chloroform	7.37	83	24645	2.059 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
Data File : VN050864.D
Acq On : 24 Aug 2018 9:37
Operator : MD\SY
Sample : J4635-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1

Quant Time: Aug 24 14:16:40 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

