

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
 Data File : VN050876.D
 Acq On : 24 Aug 2018 14:53
 Operator : MD\SY
 Sample : PB112339ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112339ZHE#02

Quant Time: Aug 24 23:35:25 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	638700	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1044600	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	885835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	289926	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	409286	53.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.46%	
35) Dibromofluoromethane	7.59	113	389556	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
50) Toluene-d8	10.09	98	1421219	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
62) 4-Bromofluorobenzene	12.40	95	373700	36.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.38%	
Target Compounds						
18) Methyl Acetate	4.33	43	12908	1.778	ug/l	97
43) Isopropyl Acetate	8.17	43	318910	26.194	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
Data File : VN050876.D
Acq On : 24 Aug 2018 14:53
Operator : MD\SY
Sample : PB112339ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112339ZHE#02

Quant Time: Aug 24 23:35:25 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

