

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
 Data File : VN050869.D
 Acq On : 24 Aug 2018 12:00
 Operator : MD\SY
 Sample : VN0824MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Manual Integrations
 APPROVED

MMDadoda
 8/25/2018 1:18:38 PM

Quant Time: Aug 24 23:08:49 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	596254	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	863287	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	776253	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	380314	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	348503	48.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.10%	
35) Dibromofluoromethane	7.59	113	333942	48.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.24%	
50) Toluene-d8	10.09	98	1249566	48.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.32%	
62) 4-Bromofluorobenzene	12.40	95	402927	47.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.44%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.85	85	169973	21.301	ug/l	99
3) Chloromethane	2.06	50	189332	19.237	ug/l	99
4) Vinyl Chloride	2.18	62	183375	19.669	ug/l	100
5) Bromomethane	2.57	94	106264	19.400	ug/l	96
6) Chloroethane	2.70	64	105501	19.585	ug/l	99
7) Trichlorofluoromethane	3.01	101	237493	19.893	ug/l	99
8) Diethyl Ether	3.41	74	80582	19.851	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	3.76	101	146466	19.637	ug/l	99
10) Methyl Iodide	3.95	142	106680	20.572	ug/l	99
11) Tert butyl alcohol	4.79	59	44990	98.428	ug/l #	86
12) 1,1-Dichloroethene	3.73	96	131525	19.640	ug/l	99
13) Acrolein	3.61	56	19600	86.084	ug/l	98
14) Allyl chloride	4.33	41	204412	19.589	ug/l	99
15) Acrylonitrile	4.99	53	227073	97.752	ug/l	100
16) Acetone	3.82	43	170041	95.676	ug/l	99
17) Carbon Disulfide	4.05	76	397792	17.719	ug/l	99
18) Methyl Acetate	4.33	43	134940	19.908	ug/l	100
19) Methyl tert-butyl Ether	5.05	73	343470	20.390	ug/l	99
20) Methylene Chloride	4.55	84	152853	18.894	ug/l	98
21) trans-1,2-Dichloroethene	5.05	96	137948	19.045	ug/l	99
22) Diisopropyl ether	5.95	45	430759	20.713	ug/l	96
23) Vinyl Acetate	5.90	43	1317231	98.354	ug/l	100
24) 1,1-Dichloroethane	5.85	63	262186	19.638	ug/l	98
25) 2-Butanone	6.84	43	251315	86.879	ug/l	99
26) 2,2-Dichloropropane	6.83	77	210363	19.609	ug/l	100
27) cis-1,2-Dichloroethene	6.83	96	156749	20.073	ug/l	99
28) Bromochloromethane	7.20	49	109888	19.370	ug/l	99
29) Tetrahydrofuran	7.22	42	168456	101.570	ug/l	99
30) Chloroform	7.37	83	264823	20.031	ug/l	100
31) Cyclohexane	7.65	56	232126	19.415	ug/l	97
32) 1,1,1-Trichloroethane	7.57	97	221533	19.508	ug/l	99
36) 1,1-Dichloropropene	7.80	75	196617	19.300	ug/l	98
37) Ethyl Acetate	6.93	43	116118	19.843	ug/l	99
38) Carbon Tetrachloride	7.78	117	201411	19.559	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN082418\
Data File : VN050869.D
Acq On : 24 Aug 2018 12:00
Operator : MD\SY
Sample : VN0824MBS01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBS01


Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM

Quant Time: Aug 24 23:08:49 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.91	180	103289	17.095	ug/l	99
94) Hexachlorobutadiene	15.01	225	84970	19.780	ug/l	100
95) Naphthalene	15.13	128	183933	16.197	ug/l	99
96) 1,2,3-Trichlorobenzene	15.32	180	105001	17.650	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampled :

VN0824MBS01

Tgt Ion:168 Resp: 596254

Ion Ratio Lower Upper

168 100





99 49.7 39.6 59.4

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#2

Dichlorodifluoromethane

Concen: 21.301 ug/l

RT: 1.85 min Scan# 81

Delta R.T. 0.00 min

Lab File: VN050869.D




Acq: 24 Aug 2018 12:00


Tgt Ion: 85 Resp: 169973

Ion Ratio Lower Upper

85 100

87 33.1 16.4 49.1





#3

Chloromethane

Concen: 19.237 ug/l

RT: 2.06 min Scan# 145

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 50 Resp: 189332

Ion Ratio Lower Upper

50 100





52 33.5 26.3 39.5

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#4

Vinyl Chloride

Concen: 19.669 ug/l

RT: 2.18 min Scan# 184

Delta R.T. 0.00 min

Lab File: VN050869.D




Acq: 24 Aug 2018 12:00


Tgt Ion: 62 Resp: 183375

Ion Ratio Lower Upper

62 100

64 31.8 25.4 38.2





#5

Bromomethane

Concen: 19.400 ug/l

RT: 2.57 min Scan# 304

Delta R.T. 0.01 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 94 Resp: 106264

Ion Ratio Lower Upper

94 100


96 91.5 76.2 114.2

Manual Integrations

APPROVED




MMDadoda

8/25/2018 1:18:38 PM



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0



#6

Chloroethane

Concen: 19.585 ug/l

RT: 2.70 min Scan# 346

Delta R.T. 0.00 min

Lab File: VN050869.D


Acq: 24 Aug 2018 12:00

Tgt Ion: 64 Resp: 105501

Ion Ratio Lower Upper



64 100


66 31.2 25.3 37.9



Abundance Ion 63.90 (63.60 to 64.60): VN0

Ion 65.90 (65.60 to 66.60): VN0





#7





Trichlorofluoromethane
 Concen: 19.893 ug/l
 RT: 3.01 min Scan# 443
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
101	100	237493		
103	65.3	51.6	77.4	

Manual Integrations
 APPROVED




MMDadoda
 8/25/2018 1:18:38 PM




#8

Diethyl Ether
 Concen: 19.851 ug/l
 RT: 3.41 min Scan# 567
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
74	100	80582		
45	95.6	48.2	144.6	





#9

1,1,2-Trichlorotrifluoroethane

Concen: 19.637 ug/l

RT: 3.76 min Scan# 675

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampleId :

VN0824MBS01

Tgt Ion:	101	Resp:	146466
Ion Ratio	Lower	Upper	
101	100		
85	42.7	33.7	50.5
151	84.8	67.4	101.2

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#10

Methyl Iodide

Concen: 20.572 ug/l




RT: 3.95 min Scan# 734


Delta R.T. -0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	142	Resp:	106680
Ion Ratio	Lower	Upper	
142	100		
127	40.4	31.7	47.5
141	13.9	11.4	17.0









#11
Tert butyl alcohol
Concen: 98.428 ug/l
RT: 4.79 min Scan# 996
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
59	100	44990		
57	5.1	8.3	12.5#	




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#12
1,1-Dichloroethene
Concen: 19.640 ug/l
RT: 3.73 min Scan# 667
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
96	100	131525		
61	154.7	124.9	187.3	
98	63.4	51.1	76.7	





#13

Acrolein

Concen: 86.084 ug/l

RT: 3.61 min Scan# 628

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 56 Resp: 19600

Ion Ratio Lower Upper

56 100





55 70.6 57.7 86.5

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#14

Allyl chloride

Concen: 19.589 ug/l

RT: 4.33 min Scan# 851

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00




Tgt Ion: 41 Resp: 204412


Ion Ratio Lower Upper

41 100

39 61.4 50.2 75.2

76 37.8 29.8 44.8





#15

Acrylonitrile

Concen: 97.752 ug/l

RT: 4.99 min Scan# 1058

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampleId :

VN0824MBS01

Tgt Ion:	53	Resp:	227073
Ion	Ratio	Lower	Upper
53	100		
52	82.3	65.9	98.9
51	35.8	28.6	43.0

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#16

Acetone

Concen: 95.676 ug/l




RT: 3.82 min Scan# 694


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	43	Resp:	170041
Ion	Ratio	Lower	Upper
43	100		
58	34.5	27.2	40.8





#17

Carbon Disulfide

Concen: 17.719 ug/l

RT: 4.05 min Scan# 766

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 76 Resp: 397792

Ion Ratio Lower Upper

76 100


78 8.9 7.3 10.9

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0

150000

100000



50000

0

4.05

4.00 4.10 4.20

Time-->



#18

Methyl Acetate

Concen: 19.908 ug/l

RT: 4.33 min Scan# 852

Delta R.T. 0.00 min

Lab File: VN050869.D


Acq: 24 Aug 2018 12:00

Tgt Ion: 43 Resp: 134940

Ion Ratio Lower Upper

43 100

74 24.8 19.8 29.8



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0

50000

40000

30000

20000


10000


0

4.33

4.20 4.30 4.40

Time-->









#19
Methyl tert-butyl Ether
Concen: 20.390 ug/l
RT: 5.05 min Scan# 1076
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampled :
VN0824MBS01

Tgt Ion: 73 Resp: 343470
Ion Ratio Lower Upper
73 100
57 22.9 17.8 26.8




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#20
Methylene Chloride
Concen: 18.894 ug/l
RT: 4.55 min Scan# 922
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion: 84 Resp: 152853
Ion Ratio Lower Upper
84 100
49 118.8 93.2 139.8
51 36.4 29.0 43.6
86 64.9 49.8 74.8





#21

trans-1,2-Dichloroethene

Concen: 19.045 ug/l

RT: 5.05 min Scan# 1075

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 96 Resp: 137948

Ion Ratio Lower Upper

96 100

61 136.4 107.4 161.0


98 63.9 50.7 76.1

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM






Abundance

Ion 95.90 (95.60 to 96.60): VNC

Ion 60.90 (60.60 to 61.60): VNC

Ion 97.90 (97.60 to 98.60): VNC



#22

Diisopropyl ether

Concen: 20.713 ug/l

RT: 5.95 min Scan# 1357

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion: 45 Resp: 430759

Ion Ratio Lower Upper

45 100

43 61.6 45.9 68.9

87 29.6 22.7 34.1

59 12.4 9.3 13.9




Abundance


Ion 45.00 (44.70 to 45.70): VNC

Ion 43.00 (42.70 to 43.70): VNC

Ion 87.00 (86.70 to 87.70): VNC

Ion 59.00 (58.70 to 59.70): VNC









#23
 Vinyl Acetate
 Concen: 98.354 ug/l
 RT: 5.90 min Scan# 1340
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion: 43 Resp: 1317231
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.7 13.1




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#24
 1,1-Dichloroethane
 Concen: 19.638 ug/l
 RT: 5.85 min Scan# 1325
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion: 63 Resp: 262186
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.3 9.8
 100 4.6 2.1 6.5





#25

2-Butanone

Concen: 86.879 ug/l

RT: 6.84 min Scan# 1633

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 43 Resp: 251315

Ion Ratio Lower Upper

43 100





72 27.6 22.3 33.5

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#26

2,2-Dichloropropane

Concen: 19.609 ug/l

RT: 6.83 min Scan# 1629

Delta R.T. 0.00 min

Lab File: VN050869.D




Acq: 24 Aug 2018 12:00


Tgt Ion: 77 Resp: 210363

Ion Ratio Lower Upper

77 100

97 24.2 12.0 36.0





#27

cis-1,2-Dichloroethene

Concen: 20.073 ug/l

RT: 6.83 min Scan# 1631

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampleId :

VN0824MBS01

Tgt Ion:	96	Resp:	156749
Ion	Ratio	Lower	Upper
96	100		
61	140.5	0.0	283.4
98	64.9	0.0	129.8

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#28

Bromochloromethane

Concen: 19.370 ug/l




RT: 7.20 min Scan# 1744


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	49	Resp:	109888
Ion	Ratio	Lower	Upper
49	100		
129	0.5	0.0	4.2
130	82.9	67.1	100.7





#29

Tetrahydrofuran

Concen: 101.570 ug/l

RT: 7.22 min Scan# 1750

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampleId :

VN0824MBS01

Tgt Ion:	42	Resp:	168456
Ion Ratio		Lower	Upper
42	100		
72	43.6	35.5	53.3
71	42.7	33.7	50.5

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#30

Chloroform

Concen: 20.031 ug/l




RT: 7.37 min Scan# 1799


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	83	Resp:	264823
Ion Ratio		Lower	Upper
83	100		
85	65.5	52.2	78.4





#31

Cyclohexane

Concen: 19.415 ug/l

RT: 7.65 min Scan# 1886

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampled :

VN0824MBS01

Tgt Ion: 56 Resp: 232126
Ion Ratio Lower Upper


56 100

69 35.2 26.2 39.4




84 88.3 68.3 102.5

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0



#32

1,1,1-Trichloroethane

Concen: 19.508 ug/l

RT: 7.57 min Scan# 1861

Delta R.T. 0.00 min

Lab File: VN050869.D


Acq: 24 Aug 2018 12:00

Tgt Ion: 97 Resp: 221533
Ion Ratio Lower Upper



97 100


99 65.8 51.8 77.6

61 43.1 34.9 52.3



Abundance Ion 96.90 (96.60 to 97.60): VN0
Ion 98.90 (98.60 to 99.60): VN0
Ion 60.90 (60.60 to 61.60): VN0









#33
 1,2-Dichloroethane-d4
 Concen: 48.551 ug/l
 RT: 8.03 min Scan# 2003
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
65	100	348503		
67	55.6	0.0	110.4	




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2176
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
114	100	863287		
63	19.7	0.0	38.6	
88	15.5	0.0	30.6	









#35
 Dibromofluoromethane
 Concen: 48.623 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
113	100			
111	102.3	82.6	123.8	
192	22.3	18.0	27.0	




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#36
 1,1-Dichloropropene
 Concen: 19.300 ug/l
 RT: 7.80 min Scan# 1930
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
75	100			
110	37.4	18.1	54.4	
77	31.7	24.8	37.2	









#37
Ethyl Acetate
 Concen: 19.843 ug/l
 RT: 6.93 min Scan# 1662
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampled :
 VN0824MBS01

Tgt Ion	Ratio	Lower	Upper
43	100		
61	15.6	12.0	18.0
70	10.7	8.6	13.0




Manual Integrations
APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#38
Carbon Tetrachloride
 Concen: 19.559 ug/l
 RT: 7.78 min Scan# 1924
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Lower	Upper
117	100		
119	96.5	76.3	114.5
121	30.9	24.6	36.8









#39
Methylcyclohexane
Concen: 18.806 ug/l
RT: 9.08 min Scan# 2330
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampled :
VN0824MBS01

Tgt Ion	Ratio	Resp Lower	Resp Upper
83	100		
55	78.0	61.0	91.4
98	46.0	37.5	56.3




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#40
Benzene
Concen: 19.891 ug/l
RT: 8.04 min Scan# 2007
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp Lower	Resp Upper
78	100		
77	23.8	19.0	28.4









#41
Methacrylonitrile
Concen: 20.743 ug/l
RT: 7.18 min Scan# 1738
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBS01

Tgt Ion	Ratio	Lower	Upper
41	100		
39	48.1	45.4	68.0
67	78.5	70.6	106.0
52	32.5	27.1	40.7




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#42
1,2-Dichloroethane
Concen: 19.726 ug/l
RT: 8.13 min Scan# 2033
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Lower	Upper
62	100		
98	10.0	0.0	19.6





#43

Isopropyl Acetate

Concen: 20.069 ug/l

RT: 8.17 min Scan# 2045

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampleId :

VN0824MBS01

Tgt Ion:	43	Resp:	201927
Ion	Ratio	Lower	Upper
43	100		
61	20.8	17.5	26.3
87	13.9	11.3	16.9

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#44

Trichloroethene

Concen: 19.591 ug/l




RT: 8.84 min Scan# 2254


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	130	Resp:	158718
Ion	Ratio	Lower	Upper
130	100		
95	96.3	0.0	193.0









#45
 1,2-Dichloropropane
 Concen: 19.850 ug/l
 RT: 9.12 min Scan# 2342
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampled :
 VN0824MBS01

Tgt Ion: 63 Resp: 160265
 Ion Ratio Lower Upper
 63 100
 65 31.3 25.0 37.6




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#46
 Dibromomethane
 Concen: 19.678 ug/l
 RT: 9.21 min Scan# 2370
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion: 93 Resp: 93350
 Ion Ratio Lower Upper
 93 100
 95 84.3 66.7 100.1
 174 112.5 89.8 134.6









#47
 Bromodichloromethane
 Concen: 20.105 ug/l
 RT: 9.40 min Scan# 2430
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
83	100	199767		
85	64.1	52.6	78.8	
127	8.7	7.2	10.8	




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#48
 Methyl methacrylate
 Concen: 19.021 ug/l
 RT: 9.20 min Scan# 2367
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
41	100	97536		
69	93.0	74.7	112.1	
39	54.0	42.2	63.2	





#49

1,4-Dioxane

Concen: 394.406 ug/l

RT: 9.20 min Scan# 2367

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N


ClientSampled :

VN0824MBS01

Tgt Ion:	88	Resp:	27570
Ion	Ratio	Lower	Upper
88	100		
43	31.8	25.9	38.9
58	73.6	55.0	82.4

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM






Abundance

Ion 88.00 (87.70 to 88.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 48.664 ug/l


RT: 10.09 min Scan# 2644

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00



Tgt Ion:	98	Resp:	1249566
Ion	Ratio	Lower	Upper
98	100		
100	63.7	51.5	77.3




Abundance

Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0









#51
 4-Methyl-2-Pentanone
 Concen: 94.372 ug/l
 RT: 9.99 min Scan# 2611
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion: 43 Resp: 566631
 Ion Ratio Lower Upper
 43 100
 58 40.0 32.4 48.6




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#52
 Toluene
 Concen: 20.326 ug/l
 RT: 10.16 min Scan# 2664
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion: 92 Resp: 369885
 Ion Ratio Lower Upper
 92 100
 91 175.5 140.6 211.0





#53

t-1,3-Dichloropropene

Concen: 18.740 ug/l

RT: 10.38 min Scan# 2735

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampled :

VN0824MBS01

Tgt Ion: 75 Resp: 185339

Ion Ratio Lower Upper

75 100





77 32.3 25.7 38.5

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#54

cis-1,3-Dichloropropene

Concen: 19.546 ug/l

RT: 9.84 min Scan# 2566

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00




Tgt Ion: 75 Resp: 221363


Ion Ratio Lower Upper

75 100

77 32.2 25.4 38.2

39 42.5 33.0 49.4









#55
 1,1,2-Trichloroethane
 Concen: 19.884 ug/l
 RT: 10.56 min Scan# 2791
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.3	69.7	104.5
85	56.9	44.5	66.7
99	61.2	50.7	76.1




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#56
 Ethyl methacrylate
 Concen: 18.741 ug/l
 RT: 10.43 min Scan# 2749
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Lower	Upper
69	100		
41	62.2	48.3	72.5
39	29.4	24.0	36.0









#57
 1,3-Dichloropropane
 Concen: 19.910 ug/l
 RT: 10.71 min Scan# 2836
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampled :
 VN0824MBS01

Tgt Ion: 76 Resp: 221962
 Ion Ratio Lower Upper
 76 100
 78 31.7 25.5 38.3




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#58
 2-Chloroethyl Vinyl ether
 Concen: 91.206 ug/l
 RT: 9.70 min Scan# 2521
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion: 63 Resp: 318389
 Ion Ratio Lower Upper
 63 100
 106 27.5 21.9 32.9









#59
2-Hexanone
Concen: 90.496 ug/l
RT: 10.75 min Scan# 2850
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBS01

Tgt Ion: 43 Resp: 361727
Ion Ratio Lower Upper
43 100
58 55.5 28.3 85.0




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#60
Dibromochloromethane
Concen: 19.400 ug/l
RT: 10.90 min Scan# 2896
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion: 129 Resp: 146708
Ion Ratio Lower Upper
129 100
127 77.8 38.3 114.9





#61

1,2-Dibromoethane

Concen: 19.925 ug/l

RT: 11.01 min Scan# 2929

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 107 Resp: 130133





Ion Ratio Lower Upper

107 100

109 93.5 75.4 113.2

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#62

4-Bromofluorobenzene

Concen: 47.222 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00




Tgt Ion: 95 Resp: 402927


Ion Ratio Lower Upper

95 100

174 89.0 0.0 183.0

176 87.0 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampled :
VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	776253		
82	53.0	42.6	64.0	
119	32.0	25.8	38.6	




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#64
Tetrachloroethene
Concen: 19.517 ug/l
RT: 10.63 min Scan# 2812
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
164	100	143897		
166	125.5	103.5	155.3	
129	91.1	72.7	109.1	
131	86.8	69.5	104.3	









#65
Chlorobenzene
Concen: 19.581 ug/l
RT: 11.43 min Scan# 3061
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampled :
VN0824MBS01

Tgt Ion:112 Resp: 397300
Ion Ratio Lower Upper
112 100
114 32.1 25.6 38.4




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#66
1,1,1,2-Tetrachloroethane
Concen: 19.819 ug/l
RT: 11.51 min Scan# 3085
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion:131 Resp: 147129
Ion Ratio Lower Upper
131 100
133 96.9 47.4 142.3
119 67.0 32.8 98.4





#67

Ethyl Benzene

Concen: 19.739 ug/l

RT: 11.51 min Scan# 3086

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampled :

VN0824MBS01

Tgt Ion: 91 Resp: 651107

Ion Ratio Lower Upper

91 100





106 31.5 25.0 37.4

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#68

m/p-Xylenes

Concen: 40.819 ug/l

RT: 11.62 min Scan# 3120

Delta R.T. 0.00 min

Lab File: VN050869.D




Acq: 24 Aug 2018 12:00


Tgt Ion: 106 Resp: 511070

Ion Ratio Lower Upper

106 100

91 202.2 161.0 241.6









#69
o-Xylene
Concen: 20.371 ug/l
RT: 11.95 min Scan# 3222
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
106	100	245210		
91	212.2	106.5	319.5	




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#70
Styrene
Concen: 18.940 ug/l
RT: 11.96 min Scan# 3226
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
104	100	384597		
78	47.7	38.2	57.4	
103	55.5	44.4	66.6	





#71

Bromoform

Concen: 19.594 ug/l

RT: 12.13 min Scan# 3277

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampleId :

VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
173	100	100032		
175	48.6	24.4	73.4	
254	0.0	0.1	0.1#	

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.34 min Scan# 3655


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	380314		
115	56.5	27.8	83.4	
150	164.9	0.0	351.4	









#73
Isopropylbenzene
Concen: 20.388 ug/l
RT: 12.25 min Scan# 3315
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBS01

Tgt Ion: 105 Resp: 642045
Ion Ratio Lower Upper
105 100
120 27.0 13.5 40.4




Manual Integrations
APPROVED


MMDadoda
8/25/2018 1:18:38 PM



#74
N-ethyl acetate
Concen: 19.126 ug/l
RT: 12.07 min Scan# 3259
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion: 43 Resp: 151337
Ion Ratio Lower Upper
43 100
70 44.3 36.2 54.2
55 27.8 21.8 32.8
61 24.9 20.5 30.7





#75

1,1,2,2-Tetrachloroethane

Concen: 20.884 ug/l

RT: 12.50 min Scan# 3394

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampled :

VN0824MBS01

Tgt Ion: 83 Resp: 163946
Ion Ratio Lower Upper


83 100

131 10.8 5.3 15.9




85 64.3 31.9 95.7

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



Abundance Ion 82.90 (82.60 to 83.60): VN0
Ion 130.80 (130.50 to 131.50): V
Ion 84.90 (84.60 to 85.60): VN0



#76

1,2,3-Trichloropropane

Concen: 21.045 ug/l m

RT: 12.55 min Scan# 3410

Delta R.T. 0.00 min


Lab File: VN050869.D

Acq: 24 Aug 2018 12:00



Tgt Ion: 75 Resp: 140188
Ion Ratio Lower Upper


75 100

77 0.0 0.0 0.0



Abundance Ion 74.90 (74.60 to 75.60): VN0
Ion 77.00 (76.70 to 77.70): VN0





#77

Bromobenzene

Concen: 19.662 ug/l

RT: 12.53 min Scan# 3401

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N


ClientSampleId :

VN0824MBS01

Tgt Ion:	156	Resp:	168278
Ion Ratio	Lower	Upper	
156	100		
77	176.6	89.9	269.7
158	96.6	49.3	147.9

Manual Integrations
APPROVED




MMDadoda
8/25/2018 1:18:38 PM



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VN0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 20.541 ug/l


RT: 12.59 min Scan# 3422

Delta R.T. 0.00 min

Lab File: VN050869.D



Acq: 24 Aug 2018 12:00


Tgt Ion:	91	Resp:	722610
Ion Ratio	Lower	Upper	
91	100		
120	23.5	11.7	35.0



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 20.227 ug/l

RT: 12.67 min Scan# 3447

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion: 91 Resp: 443233

Ion Ratio Lower Upper

91 100





126 35.9 17.8 53.3

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#80

1,3,5-Trimethylbenzene

Concen: 21.068 ug/l

RT: 12.73 min Scan# 3465

Delta R.T. 0.00 min

Lab File: VN050869.D




Acq: 24 Aug 2018 12:00


Tgt Ion: 105 Resp: 535348

Ion Ratio Lower Upper

105 100

120 49.3 24.6 74.0





#81

trans-1,4-Dichloro-2-butene

Concen: 19.497 ug/l

RT: 12.30 min Scan# 3331

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampled :

VN0824MBS01

Tgt Ion:	75	Resp:	38418
Ion	Ratio	Lower	Upper
75	100		
53	89.5	68.9	103.3
89	43.7	36.1	54.1

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#82

4-Chlorotoluene

Concen: 20.334 ug/l




RT: 12.77 min Scan# 3478


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	91	Resp:	439880
Ion	Ratio	Lower	Upper
91	100		
126	34.5	17.5	52.5









#83
 tert-Butylbenzene
 Concen: 20.567 ug/l
 RT: 12.99 min Scan# 3547
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampled :
 VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
119	100	452593		
91	63.8	31.6	94.7	
134	27.0	13.4	40.1	




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#84
 1,2,4-Trimethylbenzene
 Concen: 20.709 ug/l
 RT: 13.04 min Scan# 3561
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
105	100	529326		
120	46.5	22.9	68.8	





#85

sec-Butylbenzene

Concen: 20.722 ug/l

RT: 13.17 min Scan# 3602

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N

ClientSampleId :

VN0824MBS01

Tgt Ion:105 Resp: 597548

Ion Ratio Lower Upper

105 100





134 20.0 10.3 30.8

Manual Integrations

APPROVED

MMDadoda

8/25/2018 1:18:38 PM



#86

p-Isopropyltoluene

Concen: 21.071 ug/l

RT: 13.29 min Scan# 3639

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00




Tgt Ion:119 Resp: 517722


Ion Ratio Lower Upper

119 100

134 26.5 13.7 41.0

91 22.1 11.2 33.5









#87
 1,3-Dichlorobenzene
 Concen: 19.471 ug/l
 RT: 13.28 min Scan# 3637
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	282800		
111	38.4	19.1	57.3	
148	64.3	32.2	96.6	




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#88
 1,4-Dichlorobenzene
 Concen: 19.029 ug/l
 RT: 13.36 min Scan# 3662
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	271997		
111	38.7	18.8	56.3	
148	64.1	32.2	96.6	





#89

n-Butylbenzene

Concen: 18.967 ug/l

RT: 13.62 min Scan# 3740

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N





ClientSampleId :

VN0824MBS01

Tgt Ion:	91	Resp:	372740
Ion	Ratio	Lower	Upper
91	100		
92	51.6	26.1	78.2
134	25.9	13.2	39.6

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#90

Hexachloroethane

Concen: 20.147 ug/l




RT: 13.87 min Scan# 3820


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	117	Resp:	93828
Ion	Ratio	Lower	Upper
117	100		
201	94.7	46.0	137.8









#91
 1,2-Dichlorobenzene
 Concen: 19.990 ug/l
 RT: 13.65 min Scan# 3752
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	280101		
111	39.2	19.8	59.3	
148	64.3	32.3	96.8	




Manual Integrations
 APPROVED


MMDadoda
 8/25/2018 1:18:38 PM



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 19.255 ug/l
 RT: 14.27 min Scan# 3943
 Delta R.T. 0.00 min
 Lab File: VN050869.D
 Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
75	100	21777		
155	91.2	46.4	139.1	
157	120.4	60.7	182.0	





#93

1,2,4-Trichlorobenzene

Concen: 17.095 ug/l

RT: 14.91 min Scan# 4142

Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Instrument :

MSVOA_N




ClientSampled :

VN0824MBS01

Tgt Ion:	180	Resp:	103289
Ion Ratio	Lower	Upper	
180	100		
182	94.8	48.2	144.6
145	28.4	14.1	42.2

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#94

Hexachlorobutadiene

Concen: 19.780 ug/l




RT: 15.01 min Scan# 4174


Delta R.T. 0.00 min

Lab File: VN050869.D

Acq: 24 Aug 2018 12:00

Tgt Ion:	225	Resp:	84970
Ion Ratio	Lower	Upper	
225	100		
223	62.4	31.3	93.8
227	64.8	32.0	96.2









#95
Naphthalene
Concen: 16.197 ug/l
RT: 15.13 min Scan# 4212
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBS01

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	183933		
127	12.8	10.5	15.7	
129	10.7	8.7	13.1	

Manual Integrations
APPROVED

MMDadoda
8/25/2018 1:18:38 PM



#96
1,2,3-Trichlorobenzene
Concen: 17.650 ug/l
RT: 15.32 min Scan# 4269
Delta R.T. 0.00 min
Lab File: VN050869.D
Acq: 24 Aug 2018 12:00

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	105001		
182	96.7	48.0	144.2	
145	30.1	15.2	45.6	

