


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082418\
Data File : VN050860.D
Acq On : 24 Aug 2018 7:30
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Title : SW846 8260
Last Update : Wed Aug 22 05:38:12 2018



AutoFind: Scans 3361, 3362, 3363; Background Corrected with Scan 3348

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.6	10271	PASS
75	95	30	60	45.7	28357	PASS
95	95	100	100	100.0	62000	PASS
96	95	5	9	7.2	4447	PASS
173	174	0.00	2	1.6	877	PASS
174	95	50	100	90.9	56330	PASS
175	174	5	9	7.4	4162	PASS
176	174	95	101	98.6	55533	PASS
177	176	5	9	6.6	3660	PASS