

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082419\
 Data File : VN057480.D
 Acq On : 23 Aug 2019 18:45
 Operator : JC/SP
 Sample : K4472-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-S2(2.5)

Quant Time: Aug 26 05:35:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	407487	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	707036	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	577522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	138784	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	338307	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
35) Dibromofluoromethane	7.58	113	228473	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
50) Toluene-d8	10.08	98	885810	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
62) 4-Bromofluorobenzene	12.40	95	243433	39.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.56%	
Target Compounds						
16) Acetone	3.80	43	30752	9.342	ug/l	96
20) Methylene Chloride	4.53	84	15296	3.064	ug/l	95
43) Isopropyl Acetate	8.16	43	82016	6.266	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082419\
Data File : VN057480.D
Acq On : 23 Aug 2019 18:45
Operator : JC/SP
Sample : K4472-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-S2(2.5)

Quant Time: Aug 26 05:35:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

