

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN082419\
 Data File : VN057496.D
 Acq On : 24 Aug 2019 1:34
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 06:40:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.590	0.572	3.1	79	0.00
3 P	Chloromethane	0.843	0.788	6.5	82	0.00
4 C	Vinyl Chloride	1.026	1.005	2.0#	83	0.00
5 T	Bromomethane	0.632	0.618	2.2	86	0.00
6 T	Chloroethane	0.660	0.657	0.5	88	0.00
7 T	Trichlorofluoromethane	1.592	1.567	1.6	87	0.00
8 T	Diethyl Ether	0.601	0.564	6.2	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.881	0.826	6.2	86	0.00
10 T	Methyl Iodide	0.650	0.642	1.2	81	0.00
11 T	Tert butyl alcohol	0.109	0.098	10.1	78	0.00
12 CM	1,1-Dichloroethene	0.851	0.838	1.5#	87	0.00
13 T	Acrolein	0.096	0.092	4.2	79	0.00
14 T	Allyl chloride	1.231	1.083	12.0	80	0.00
15 T	Acrylonitrile	0.295	0.289	2.0	83	0.00
16 T	Acetone	0.404	0.395	2.2	124	0.00
17 T	Carbon Disulfide	1.623	1.439	11.3	80	0.00
18 T	Methyl Acetate	1.047	0.898	14.2	85	0.00
19 T	Methyl tert-butyl Ether	2.054	1.905	7.3	82	0.00
20 T	Methylene Chloride	0.660	0.596	9.7	83	0.00
21 T	trans-1,2-Dichloroethene	0.601	0.555	7.7	84	0.00
22 T	Diisopropyl ether	2.415	2.385	1.2	85	0.00
23 T	Vinyl Acetate	1.886	1.786	5.3	81	0.00
24 P	1,1-Dichloroethane	1.280	1.260	1.6	87	0.00
25 T	2-Butanone	0.439	0.406	7.5	81	0.00
26 T	2,2-Dichloropropane	1.108	0.806	27.3#	65	0.00
27 T	cis-1,2-Dichloroethene	0.712	0.684	3.9	86	0.00
28 T	Bromochloromethane	0.689	0.748	-8.6	87	0.00
29 T	Tetrahydrofuran	0.289	0.275	4.8	82	0.00
30 C	Chloroform	1.328	1.216	8.4#	86	0.00
31 T	Cyclohexane	1.236	1.201	2.8	84	0.00
32 T	1,1,1-Trichloroethane	1.098	1.024	6.7	82	0.00
33 S	1,2-Dichloroethane-d4	0.880	0.943	-7.2	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.338	0.329	2.7	84	0.00
36 T	1,1-Dichloropropene	0.511	0.481	5.9	86	0.00
37 T	Ethyl Acetate	0.505	0.473	6.3	84	0.00
38 T	Carbon Tetrachloride	0.508	0.442	13.0	79	0.00
39 T	Methylcyclohexane	0.607	0.567	6.6	83	0.00
40 TM	Benzene	1.533	1.424	7.1	85	0.00
41 T	Methacrylonitrile	0.241	0.232	3.7	90	0.00
42 TM	1,2-Dichloroethane	0.590	0.560	5.1	88	0.00
43 T	Isopropyl Acetate	0.926	0.845	8.7	82	0.00
44 TM	Trichloroethene	0.342	0.316	7.6	82	0.00
45 C	1,2-Dichloropropane	0.432	0.411	4.9#	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN082419\
 Data File : VN057496.D
 Acq On : 24 Aug 2019 1:34
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 26 06:40:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 23 05:07:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.276	0.240	13.0	82	0.00
47 T	Bromodichloromethane	0.570	0.519	8.9	85	0.00
48 T	Methyl methacrylate	0.440	0.418	5.0	85	0.00
49 T	1,4-Dioxane	0.006	0.005	16.7	86	0.00
50 S	Toluene-d8	1.277	1.352	-5.9	87	0.00
51 T	4-Methyl-2-Pentanone	0.516	0.494	4.3	85	0.00
52 CM	Toluene	0.884	0.883	0.1#	89	0.00
53 T	t-1,3-Dichloropropene	0.597	0.553	7.4	85	0.00
54 T	cis-1,3-Dichloropropene	0.664	0.574	13.6	80	0.00
55 T	1,1,2-Trichloroethane	0.357	0.339	5.0	87	0.00
56 T	Ethyl methacrylate	0.562	0.551	2.0	85	0.00
57 T	1,3-Dichloropropane	0.622	0.626	-0.6	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.133	0.185	-39.1#	112	0.00
59 T	2-Hexanone	0.354	0.338	4.5	84	0.00
60 T	Dibromochloromethane	0.378	0.355	6.1	85	0.00
61 T	1,2-Dibromoethane	0.351	0.344	2.0	86	0.00
62 S	4-Bromofluorobenzene	0.433	0.466	-7.6	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.350	0.318	9.1	89	0.00
65 PM	Chlorobenzene	1.011	0.927	8.3	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.346	10.8	86	0.00
67 C	Ethyl Benzene	1.956	1.836	6.1#	88	0.00
68 T	m/p-Xylenes	0.665	0.628	5.6	86	0.00
69 T	o-Xylene	0.659	0.628	4.7	87	0.00
70 T	Styrene	1.044	1.043	0.1	88	0.00
71 P	Bromoform	0.262	0.221	15.6	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	5.589	4.498	19.5	87	0.00
74 T	N-amyl acetate	2.561	2.007	21.6	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.638	1.251	23.6	86	0.00
76 T	1,2,3-Trichloropropane	1.268	1.182	6.8	90	0.00
77 T	Bromobenzene	1.108	0.925	16.5	88	0.00
78 T	n-propylbenzene	5.927	5.030	15.1	87	0.00
79 T	2-Chlorotoluene	3.533	3.013	14.7	87	0.00
80 T	1,3,5-Trimethylbenzene	4.395	3.727	15.2	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.495	0.372	24.8	75	0.00
82 T	4-Chlorotoluene	3.368	2.889	14.2	86	0.00
83 T	tert-Butylbenzene	3.913	3.077	21.4	87	0.00
84 T	1,2,4-Trimethylbenzene	3.887	3.565	8.3	87	0.00
85 T	sec-Butylbenzene	4.787	4.152	13.3	87	0.00
86 T	p-Isopropyltoluene	3.908	3.451	11.7	86	0.00
87 T	1,3-Dichlorobenzene	1.615	1.429	11.5	85	0.00
88 T	1,4-Dichlorobenzene	1.582	1.383	12.6	84	0.00
89 T	n-Butylbenzene	3.307	2.982	9.8	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN082419\
Data File : VN057496.D
Acq On : 24 Aug 2019 1:34
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 26 06:40:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N082219W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 23 05:07:29 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.904	0.676	25.2#	81	0.00
91 T	1,2-Dichlorobenzene	1.542	1.436	6.9	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.266	0.209	21.4	81	0.00
93 T	1,2,4-Trichlorobenzene	0.562	0.443	21.2	80	0.00
94 T	Hexachlorobutadiene	0.479	0.376	21.5	85	0.00
95 T	Naphthalene	1.861	1.457	21.7	79	0.00
96 T	1,2,3-Trichlorobenzene	0.587	0.468	20.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6