

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082420\
 Data File : VN063098.D
 Acq On : 24 Aug 2020 16:16
 Operator : JC/MD
 Sample : L3780-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-082120-A

Quant Time: Aug 25 00:55:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	139773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	242460	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	250546	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97579	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	95268	57.51	ug/l	0.00
Spiked Amount			Recovery	=	115.02%	
35) Dibromofluoromethane	7.54	113	78126	49.45	ug/l	0.00
Spiked Amount			Recovery	=	98.90%	
50) Toluene-d8	10.06	98	313743	51.70	ug/l	0.00
Spiked Amount			Recovery	=	103.40%	
62) 4-Bromofluorobenzene	12.38	95	104577	48.84	ug/l	0.00
Spiked Amount			Recovery	=	97.68%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Chloroethane	2.62	64	346	0.316	ug/l #	43
11) Tert butyl alcohol	4.67	59	526	2.242	ug/l #	74
16) Acetone	3.80	43	11347	18.331	ug/l	90
18) Methyl Acetate	4.29	43	4994	3.034	ug/l	92
20) Methylene Chloride	4.48	84	11162	6.480	ug/l	95
25) 2-Butanone	6.81	43	1826	2.076	ug/l	99
29) Tetrahydrofuran	7.18	42	145	0.255	ug/l #	34
31) Cyclohexane	7.62	56	2668	1.093	ug/l #	1
36) 1,1-Dichloropropene	7.62	75	10237	4.791	ug/l #	49
38) Carbon Tetrachloride	7.62	117	12662	5.870	ug/l #	17
43) Isopropyl Acetate	8.13	43	34945	10.536	ug/l #	87
45) 1,2-Dichloropropane	9.24	63	564	0.288	ug/l #	43
48) Methyl methacrylate	9.16	41	2978	1.898	ug/l #	25
51) 4-Methyl-2-Pentanone	9.96	43	5398	2.575	ug/l	92
54) cis-1,3-Dichloropropene	9.88	75	15380	5.470	ug/l #	64
56) Ethyl methacrylate	10.51	69	1245	2.186	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	1827	0.629	ug/l #	42
59) 2-Hexanone	10.74	43	26015	17.789	ug/l #	55
71) Bromoform	12.37	173	882	0.566	ug/l #	1
74) N-amyl acetate	12.01	43	1313	0.549	ug/l #	43
76) 1,2,3-Trichloropropane	12.38	75	51619	27.215	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.38	75	51619	74.745	ug/l #	15
89) n-Butylbenzene	13.60	91	67	2.783	ug/l #	34
95) Naphthalene	15.10	128	30	5.649	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082420\
Data File : VN063098.D
Acq On : 24 Aug 2020 16:16
Operator : JC/MD
Sample : L3780-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A

Quant Time: Aug 25 00:55:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.62 min Scan# 1819

Delta R.T. -0.00 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Instrument :

MSVOA_N

ClientSampleId :


JC-03-082120-A

Tot Ion:168 Resp: 139773

Ion Ratio Lower Upper

168 100

99 53.8 42.0 63.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

60000



40000

20000

0

7.50 7.60 7.70

Time-->



#6

Chloroethane

Concen: 0.316 ug/l

RT: 2.62 min Scan# 263

Delta R.T. -0.04 min

Lab File: VN063098.D


Acq: 24 Aug 2020 16:16

Tgt Ion: 64 Resp: 346

Ion Ratio Lower Upper

64 100

66 0.0 25.4 38.0#



Abundance Ion 63.90 (63.60 to 64.60): VN0

Ion 65.90 (65.60 to 66.60): VN0

200

150


100


50

0

2.60 2.62 2.64

Time-->









#11

Tert butyl alcohol
 Concen: 2.242 ug/l
 RT: 4.67 min Scan# 901
 Delta R.T. -0.07 min
 Lab File: VN063098.D
 Acq: 24 Aug 2020 16:16

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-082120-A




Tot Ion: 59 Resp: 526
 Ion Ratio Lower Upper
 59 100
 57 0.0 7.8 11.6#




#16

Acetone
 Concen: 18.331 ug/l
 RT: 3.80 min Scan# 629
 Delta R.T. 0.02 min
 Lab File: VN063098.D
 Acq: 24 Aug 2020 16:16

Tgt Ion: 43 Resp: 11347
 Ion Ratio Lower Upper
 43 100
 58 29.0 27.8 41.6









#18
Methyl Acetate
Concen: 3.034 ug/l
RT: 4.29 min Scan# 783
Delta R.T. 0.01 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16




Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A


Tot Ion: 43 Resp: 4994
Ion Ratio Lower Upper
43 100
74 19.4 18.6 28.0



#20
Methylene Chloride
Concen: 6.480 ug/l
RT: 4.48 min Scan# 841
Delta R.T. -0.03 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Tgt Ion: 84 Resp: 11162
Ion Ratio Lower Upper
84 100
49 117.9 98.7 148.1
51 39.7 30.8 46.2
86 58.7 51.2 76.8





#25

2-Butanone

Concen: 2.076 ug/l

RT: 6.81 min Scan# 1566

Delta R.T. 0.02 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Instrument :

MSVOA_N

ClientSampleId :


JC-03-082120-A

Tot Ion: 43 Resp: 1826

Ion Ratio Lower Upper

43 100

72 27.6 21.8 32.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0



6.81

400

200

0

Time-->



#29

Tetrahydrofuran

Concen: 0.255 ug/l

RT: 7.18 min Scan# 1680

Delta R.T. 0.01 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16


Tgt Ion: 42 Resp: 145

Ion Ratio Lower Upper

42 100

72 0.0 34.6 52.0#

71 0.0 32.3 48.5#



Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0

7.18

250

200


150


100

50

0

Time-->









#31
Cyclohexane
Concen: 1.093 ug/l
RT: 7.62 min Scan# 1817
Delta R.T. 0.00 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A

Tot Ion: 56 Resp: 2668
Ion Ratio Lower Upper
56 100
69 167.4 26.6 40.0#
84 142.6 66.4 99.6#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 57.509 ug/l
RT: 7.99 min Scan# 1932
Delta R.T. -0.00 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Tgt Ion: 65 Resp: 95268
Ion Ratio Lower Upper
65 100
67 53.4 0.0 112.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2107
 Delta R.T. -0.00 min
 Lab File: VN063098.D
 Acq: 24 Aug 2020 16:16

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-082120-A

Tot Ion:114 Resp: 242460
 Ion Ratio Lower Upper
 114 100
 63 21.0 0.0 43.2
 88 15.6 0.0 30.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 49.451 ug/l
 RT: 7.54 min Scan# 1795
 Delta R.T. -0.00 min
 Lab File: VN063098.D
 Acq: 24 Aug 2020 16:16

Tgt Ion:113 Resp: 78126
 Ion Ratio Lower Upper
 113 100
 111 103.8 81.7 122.5
 192 21.8 17.5 26.3



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.791 ug/l

RT: 7.62 min Scan# 1818

Delta R.T. -0.14 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Instrument :

MSVOA_N

ClientSampleId :

JC-03-082120-A


Tot Ion: 75 Resp: 10237

Ion Ratio Lower Upper

75 100

110 7.9 17.5 52.6#




77 0.0 24.2 36.4#



Abundance Ion 74.90 (74.60 to 75.60): VN063098.D (-1767) (-)

Ion 109.90 (109.60 to 110.60): VN063098.D (-1767) (-)

Ion 76.90 (76.60 to 77.60): VN063098.D (-1767) (-)



#38

Carbon Tetrachloride

Concen: 5.870 ug/l

RT: 7.62 min Scan# 1818

Delta R.T. -0.12 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16


Tgt Ion: 117 Resp: 12662

Ion Ratio Lower Upper

117 100

119 6.0 76.2 114.4#



121 0.0 24.2 36.2#




Abundance Ion 116.80 (116.50 to 117.50): VN063098.D (-1761) (-)

Ion 118.80 (118.50 to 119.50): VN063098.D (-1761) (-)

Ion 120.80 (120.50 to 121.50): VN063098.D (-1761) (-)





#43

Isopropyl Acetate

Concen: 10.536 ug/l

RT: 8.13 min Scan# 1977

Delta R.T. 0.00 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Instrument :

MSVOA_N

ClientSampleId :

JC-03-082120-A


Tot Ion: 43 Resp: 34945

Ion Ratio Lower Upper

43 100

61 18.9 22.6 33.8#

87 11.8 10.5 15.7





Abundance Ion 43.00 (42.70 to 43.70): VN063098.D (-1957) (-)

Ion 61.00 (60.70 to 61.70): VN063098.D (-1957) (-)

Ion 87.00 (86.70 to 87.70): VN063098.D (-1957) (-)

8.13

Time-->



#45

1,2-Dichloropropane

Concen: 0.288 ug/l

RT: 9.24 min Scan# 2322

Delta R.T. 0.15 min

Lab File: VN063098.D


Acq: 24 Aug 2020 16:16

Tgt Ion: 63 Resp: 564

Ion Ratio Lower Upper

63 100

65 0.0 25.6 38.4#





Abundance Ion 62.90 (62.60 to 63.60): VN063098.D (-2181) (-)

Ion 64.90 (64.60 to 65.60): VN063098.D (-2181) (-)

9.24

Time-->









#48
Methvl methacrylate
Concen: 1.898 ug/l
RT: 9.16 min Scan# 2296
Delta R.T. -0.01 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16




Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A


Tot Ion: 41 Resp: 2978
Ion Ratio Lower Upper
41 100
69 0.0 74.2 111.4#
39 26.1 42.5 63.7#



#50
Toluene-d8
Concen: 51.702 ug/l
RT: 10.06 min Scan# 2577
Delta R.T. -0.00 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Tgt Ion: 98 Resp: 313743
Ion Ratio Lower Upper
98 100
100 64.4 51.2 76.8









#51
4-Methyl-2-Pentanone
Concen: 2.575 ug/l
RT: 9.96 min Scan# 2546
Delta R.T. 0.01 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16




Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A


Tot Ion: 43 Resp: 5398
Ion Ratio Lower Upper
43 100
58 35.3 32.2 48.2



#54
cis-1,3-Dichloropropene
Concen: 5.470 ug/l
RT: 9.88 min Scan# 2521
Delta R.T. 0.07 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Tgt Ion: 75 Resp: 15380
Ion Ratio Lower Upper
75 100
77 0.0 25.4 38.2#
39 30.7 35.7 53.5#





#56

Ethyl methacrylate

Concen: 2.186 ug/l

RT: 10.51 min Scan# 2718

Delta R.T. 0.11 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Instrument :

MSVOA_N

ClientSampleId :

JC-03-082120-A


Tot Ion: 69 Resp: 1245

Ion Ratio Lower Upper

69 100

41 2486.7 50.2 75.2#




39 1023.9 24.2 36.4#



Abundance Ion 69.00 (68.70 to 69.70): VN063098.D (-2670) (-)

Ion 41.00 (40.70 to 41.70): VN063098.D (-2670) (-)

Ion 39.00 (38.70 to 39.70): VN063098.D (-2670) (-)



#57

1,3-Dichloropropane

Concen: 0.629 ug/l

RT: 10.74 min Scan# 2788

Delta R.T. 0.06 min

Lab File: VN063098.D


Acq: 24 Aug 2020 16:16

Tgt Ion: 76 Resp: 1827

Ion Ratio Lower Upper



76 100


78 0.0 26.1 39.1#



Abundance Ion 75.90 (75.60 to 76.60): VN063098.D (-2677) (-)

Ion 77.90 (77.60 to 78.60): VN063098.D (-2677) (-)









#59
2-Hexanone
Concen: 17.789 ug/l
RT: 10.74 min Scan# 2789
Delta R.T. 0.02 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16




Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A


Tot Ion: 43 Resp: 26015
Ion Ratio Lower Upper
43 100
58 23.1 28.1 84.3#



#62
4-Bromofluorobenzene
Concen: 48.835 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. -0.00 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Tgt Ion: 95 Resp: 104577
Ion Ratio Lower Upper
95 100
174 87.2 0.0 170.8
176 83.3 0.0 164.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. -0.00 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.3	44.5	66.7
119	32.0	25.5	38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 0.566 ug/l
RT: 12.37 min Scan# 3297
Delta R.T. 0.27 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Tgt Ion	Ratio	Lower	Upper
173	100		
175	623.8	24.3	72.8#
254	0.0	0.0	0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. -0.00 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Instrument :

MSVOA_N

ClientSampleId :

JC-03-082120-A





Tot Ion:152 Resp: 97579

Ion Ratio Lower Upper

152 100

115 57.3 28.7 86.3

150 156.6 0.0 348.8



#74

N-ethyl acetate

Concen: 0.549 ug/l

RT: 12.01 min Scan# 3185

Delta R.T. -0.03 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Tgt Ion: 43 Resp: 1313




Ion Ratio Lower Upper


43 100

70 0.0 33.4 50.0#

55 0.0 20.2 30.2#

61 0.0 19.9 29.9#





#76

1,2,3-Trichloropropane

Concen: 27.215 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16

Instrument :

MSVOA_N

ClientSampleId :


JC-03-082120-A

Tot Ion: 75 Resp: 51619

Ion Ratio Lower Upper




75 100

77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#81

trans-1,4-Dichloro-2-butene

Concen: 74.745 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. 0.10 min

Lab File: VN063098.D

Acq: 24 Aug 2020 16:16


Tgt Ion: 75 Resp: 51619

Ion Ratio Lower Upper

75 100

53 0.0 69.2 103.8#



89 0.0 34.2 51.4#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0









#89
n-Butylbenzene
Concen: 2.783 ug/l
RT: 13.60 min Scan# 3679
Delta R.T. 0.01 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Instrument :
MSVOA_N
ClientSampleId :
JC-03-082120-A

Tot Ion: 91 Resp: 67
Ion Ratio Lower Upper
91 100
92 0.0 25.9 77.6#
134 0.0 12.4 37.0#



#95
Naphthalene
Concen: 5.649 ug/l
RT: 15.10 min Scan# 4146
Delta R.T. -0.00 min
Lab File: VN063098.D
Acq: 24 Aug 2020 16:16

Tgt Ion: 128 Resp: 30
Ion Ratio Lower Upper
128 100
127 0.0 10.6 16.0#
129 0.0 8.9 13.3#

