

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082421\
 Data File : VN068280.D
 Acq On : 24 Aug 2021 11:33
 Operator : JC/MD
 Sample : VN0824MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824MBL01

Quant Time: Aug 25 03:09:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	501607	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	838027	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	823106	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	292303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	290010	54.149	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.300%	
35) Dibromofluoromethane	8.024	113	263972	50.172	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.340%	
50) Toluene-d8	10.446	98	1047330	52.242	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.480%	
62) 4-Bromofluorobenzene	12.736	95	318425	47.713	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082421\
Data File : VN068280.D
Acq On : 24 Aug 2021 11:33
Operator : JC/MD
Sample : VN0824MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0824MBL01

Quant Time: Aug 25 03:09:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

