

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082421\
 Data File : VN068284.D
 Acq On : 24 Aug 2021 13:27
 Operator : JC/MD
 Sample : M3463-37ME
 Misc : 7.06g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SWCS-29-0002ME

Manual Integrations
 APPROVED

SAM
 8/25/2021 7:09:53 PM

Quant Time: Aug 25 03:10:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	491959	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.965	114	802680	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	803113	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	345057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	268720	51.158	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.320%	
35) Dibromofluoromethane	8.019	113	243305	48.280	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.560%	
50) Toluene-d8	10.440	98	995923	51.866	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.740%	
62) 4-Bromofluorobenzene	12.734	95	336595	52.657	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.320%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.319	96	8353m	1.562	ug/l	
44) Trichloroethene	9.212	130	177013	29.392	ug/l	77
64) Tetrachloroethene	10.977	164	22626	3.643	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082421\
Data File : VN068284.D
Acq On : 24 Aug 2021 13:27
Operator : JC/MD
Sample : M3463-37ME
Misc : 7.06g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SWCS-29-0002ME

Manual Integrations
APPROVED

SAM
8/25/2021 7:09:53 PM

Quant Time: Aug 25 03:10:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

