

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082421\
 Data File : VN068288.D
 Acq On : 24 Aug 2021 16:07
 Operator : JC/MD
 Sample : M3463-17
 Misc : 3.80g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SWCS-18-0002

Quant Time: Aug 25 03:11:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	441310	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.965	114	736307	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	820897	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	352450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	256330	54.400	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.800%	
35) Dibromofluoromethane	8.018	113	228705	49.474	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.940%	
50) Toluene-d8	10.440	98	935219	53.095	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.180%	
62) 4-Bromofluorobenzene	12.733	95	353489	60.284	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 120.560%	
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	8436	6.132	ug/l	99
18) Methyl Acetate	4.848	43	39122	8.454	ug/l #	89
44) Trichloroethene	9.212	130	5692	1.030	ug/l	85
52) Toluene	10.505	92	30548	2.439	ug/l	98
95) Naphthalene	15.509	128	6927	3.812	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082421\
Data File : VN068288.D
Acq On : 24 Aug 2021 16:07
Operator : JC/MD
Sample : M3463-17
Misc : 3.80g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SWCS-18-0002

Quant Time: Aug 25 03:11:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

