

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
 Data File : VN074084.D
 Acq On : 24 Aug 2022 11:56
 Operator : JC\MD
 Sample : VN0824WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0824WBL01

Quant Time: Aug 25 07:52:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	572298	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	859799	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	708782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	290377	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	288373	51.734	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.460%
35) Dibromofluoromethane	7.951	113	257665	50.674	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.340%
50) Toluene-d8	10.380	98	962784	48.890	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.780%
62) 4-Bromofluorobenzene	12.669	95	287583	44.002	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
Data File : VN074084.D
Acq On : 24 Aug 2022 11:56
Operator : JC\MD
Sample : VN0824WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0824WBL01

Quant Time: Aug 25 07:52:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

