

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
 Data File : VN074088.D
 Acq On : 24 Aug 2022 13:47
 Operator : JC\MD
 Sample : N4244-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-105I

Quant Time: Aug 25 07:53:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	616826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	928529	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.687	117	769845	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	336729	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	311276	51.811	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.620%
35) Dibromofluoromethane	7.952	113	277074	50.458	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.920%
50) Toluene-d8	10.381	98	1024181	48.158	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	12.669	95	322807	45.735	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
Data File : VN074088.D
Acq On : 24 Aug 2022 13:47
Operator : JC\MD
Sample : N4244-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-105I

Quant Time: Aug 25 07:53:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

