

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
 Data File : VN074090.D
 Acq On : 24 Aug 2022 14:36
 Operator : JC\MD
 Sample : N4244-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-104D

Quant Time: Aug 25 07:54:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	607102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	922554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	764663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	334858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	307457	51.995	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.000%
35) Dibromofluoromethane	7.951	113	274622	50.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.680%
50) Toluene-d8	10.381	98	1021268	48.332	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.660%
62) 4-Bromofluorobenzene	12.669	95	321312	45.818	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.640%
Target Compounds						
						Qvalue
3) Chloromethane	2.263	50	1836	0.343	ug/l #	79
17) Carbon Disulfide	4.446	76	6287	0.552	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
Data File : VN074090.D
Acq On : 24 Aug 2022 14:36
Operator : JC\MD
Sample : N4244-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-104D

Quant Time: Aug 25 07:54:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

