

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
 Data File : VN074099.D
 Acq On : 24 Aug 2022 18:19
 Operator : JC\MD
 Sample : N4280-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB02-20220817

Quant Time: Aug 25 07:55:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	580450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	893233	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	736617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	320714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	300274	53.112	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.220%	
35) Dibromofluoromethane	7.957	113	267549	50.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.300%	
50) Toluene-d8	10.380	98	979744	47.889	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.780%	
62) 4-Bromofluorobenzene	12.674	95	300966	44.326	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.660%	
Target Compounds						
					Qvalue	
16) Acetone	4.240	43	5137	1.926	ug/l #	76
20) Methylene Chloride	5.010	84	1763	0.313	ug/l #	80
44) Trichloroethene	9.151	130	2916	0.436	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082422\
Data File : VN074099.D
Acq On : 24 Aug 2022 18:19
Operator : JC\MD
Sample : N4280-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB02-20220817

Quant Time: Aug 25 07:55:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

